



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021021\
 Data File : VD068338.D
 Acq On : 10 Feb 2021 09:09
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D011921S.M
 Title : SW846 8260
 Last Update : Tue Jan 19 13:19:08 2021



AutoFind: Scans 1820, 1821, 1822; Background Corrected with Scan 1812

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.0	24008	PASS
75	95	30	60	53.1	67037	PASS
95	95	100	100	100.0	126275	PASS
96	95	5	9	7.0	8787	PASS
173	174	0.00	2	1.0	955	PASS
174	95	50	100	79.2	99995	PASS
175	174	5	9	7.8	7820	PASS
176	174	95	101	98.2	98181	PASS
177	176	5	9	6.5	6415	PASS