

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021122\
 Data File : VD071941.D
 Acq On : 11 Feb 2022 17:19
 Operator : VA/SY
 Sample : N1489-02
 Misc : 3.96G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 COMP-2

Quant Time: Feb 11 23:59:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 27 06:04:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	425269	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	797685	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	731663	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	297419	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	262559	48.847	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	97.700%
35) Dibromofluoromethane	7.909	113	255138	48.659	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.320%
50) Toluene-d8	10.338	98	1016185	51.893	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.780%
62) 4-Bromofluorobenzene	12.626	95	312311	45.670	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	91.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021122\
Data File : VD071941.D
Acq On : 11 Feb 2022 17:19
Operator : VA/SY
Sample : N1489-02
Misc : 3.96G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
COMP-2

Quant Time: Feb 11 23:59:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 27 06:04:38 2022
Response via : Initial Calibration

