

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021122\
 Data File : VD071928.D
 Acq On : 11 Feb 2022 08:58
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 11 23:54:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 27 06:04:38 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	0.593	0.606	-2.2	103	0.00
3 P	Chloromethane	0.710	0.653	8.0	93	0.00
4 C	Vinyl Chloride	0.743	0.731	1.6#	96	0.00
5 T	Bromomethane	0.490	0.488	0.4	102	0.00
6 T	Chloroethane	0.471	0.471	0.0	97	0.00
7 T	Trichlorofluoromethane	1.065	1.102	-3.5	100	0.00
8 T	Diethyl Ether	0.281	0.269	4.3	89	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.601	0.649	-8.0	103	0.00
10 T	Methyl Iodide	0.616	0.692	-12.3	97	0.00
11 T	Tert butyl alcohol	0.046	0.027	41.3#	66	-0.01
12 CM	1,1-Dichloroethene	0.549	0.578	-5.3#	96	0.00
13 T	Acrolein	0.056	0.042	25.0	79	0.00
14 T	Allyl chloride	1.047	1.096	-4.7	97	0.00
15 T	Acrylonitrile	0.143	0.134	6.3	85	0.00
16 T	Acetone	0.173	0.179	-3.5	90	0.00
17 T	Carbon Disulfide	1.915	1.879	1.9	93	0.00
18 T	Methyl Acetate	0.374	0.315	15.8	82	0.00
19 T	Methyl tert-butyl Ether	1.314	1.298	1.2	88	0.00
20 T	Methylene Chloride	0.713	0.632	11.4	92	0.00
21 T	trans-1,2-Dichloroethene	0.639	0.662	-3.6	96	0.00
22 T	Diisopropyl ether	2.052	2.162	-5.4	94	0.00
23 T	Vinyl Acetate	1.132	1.169	-3.3	88	0.00
24 P	1,1-Dichloroethane	1.196	1.274	-6.5	98	0.00
25 T	2-Butanone	0.202	0.191	5.4	83	0.00
26 T	2,2-Dichloropropane	0.990	1.077	-8.8	102	0.00
27 T	cis-1,2-Dichloroethene	0.687	0.729	-6.1	98	0.00
28 T	Bromochloromethane	0.451	0.415	8.0	84	0.00
29 T	Tetrahydrofuran	0.119	0.112	5.9	83	0.00
30 C	Chloroform	1.222	1.260	-3.1#	97	0.00
31 T	Cyclohexane	1.207	1.228	-1.7	100	0.00
32 T	1,1,1-Trichloroethane	1.078	1.153	-7.0	102	0.00
33 S	1,2-Dichloroethane-d4	0.632	0.570	9.8	85	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	94	0.00
35 S	Dibromofluoromethane	0.329	0.327	0.6	93	0.00
36 T	1,1-Dichloropropene	0.564	0.626	-11.0	99	0.00
37 T	Ethyl Acetate	0.262	0.250	4.6	84	0.00
38 T	Carbon Tetrachloride	0.558	0.629	-12.7	102	0.00
39 T	Methylcyclohexane	0.654	0.745	-13.9	100	0.00
40 TM	Benzene	1.523	1.675	-10.0	98	0.00
41 T	Methacrylonitrile	0.148	0.158	-6.8	87	0.00
42 TM	1,2-Dichloroethane	0.481	0.495	-2.9	91	0.00
43 T	Isopropyl Acetate	0.471	0.475	-0.8	87	0.00
44 TM	Trichloroethene	0.393	0.432	-9.9	100	0.00
45 C	1,2-Dichloropropane	0.393	0.423	-7.6#	96	0.00
46 T	Dibromomethane	0.206	0.216	-4.9	92	0.00
47 T	Bromodichloromethane	0.544	0.586	-7.7	95	0.00
48 T	Methyl methacrylate	0.249	0.246	1.2	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021122\
 Data File : VD071928.D
 Acq On : 11 Feb 2022 08:58
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 11 23:54:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 27 06:04:38 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	76	0.00
50 S	Toluene-d8	1.227	1.196	2.5	88	0.00
51 T	4-Methyl-2-Pentanone	0.247	0.245	0.8	84	0.00
52 CM	Toluene	0.947	1.058	-11.7#	99	0.00
53 T	t-1,3-Dichloropropene	0.499	0.507	-1.6	87	0.00
54 T	cis-1,3-Dichloropropene	0.585	0.602	-2.9	91	0.00
55 T	1,1,2-Trichloroethane	0.273	0.279	-2.2	90	0.00
56 T	Ethyl methacrylate	0.338	0.349	-3.3	86	0.00
57 T	1,3-Dichloropropane	0.487	0.508	-4.3	91	0.00
58 T	2-Chloroethyl Vinyl ether	0.159	0.137	13.8	73	0.00
59 T	2-Hexanone	0.179	0.181	-1.1	83	0.00
60 T	Dibromochloromethane	0.337	0.349	-3.6	91	0.00
61 T	1,2-Dibromoethane	0.263	0.268	-1.9	89	0.00
62 S	4-Bromofluorobenzene	0.429	0.418	2.6	91	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	93	0.00
64 T	Tetrachloroethene	0.354	0.391	-10.5	98	0.00
65 PM	Chlorobenzene	1.058	1.147	-8.4	96	0.00
66 T	1,1,1,2-Tetrachloroethane	0.385	0.416	-8.1	96	0.00
67 C	Ethyl Benzene	1.946	2.255	-15.9#	100	0.00
68 T	m/p-Xylenes	0.740	0.851	-15.0	98	0.00
69 T	o-Xylene	0.673	0.776	-15.3	98	0.00
70 T	Styrene	1.160	1.325	-14.2	97	0.00
71 P	Bromoform	0.205	0.211	-2.9	91	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
73 T	Isopropylbenzene	3.872	4.332	-11.9	101	0.00
74 T	N-amyl acetate	0.952	0.956	-0.4	87	0.00
75 P	1,1,2,2-Tetrachloroethane	0.716	0.679	5.2	89	0.00
76 T	1,2,3-Trichloropropane	0.527	0.490	7.0	90	0.00
77 T	Bromobenzene	0.861	0.898	-4.3	96	0.00
78 T	n-propylbenzene	4.855	5.379	-10.8	99	0.00
79 T	2-Chlorotoluene	2.751	2.963	-7.7	98	0.00
80 T	1,3,5-Trimethylbenzene	3.211	3.601	-12.1	100	0.00
81 T	trans-1,4-Dichloro-2-butene	0.214	0.184	14.0	78	0.00
82 T	4-Chlorotoluene	2.862	3.078	-7.5	98	0.00
83 T	tert-Butylbenzene	2.661	2.973	-11.7	100	0.00
84 T	1,2,4-Trimethylbenzene	3.159	3.499	-10.8	98	0.00
85 T	sec-Butylbenzene	4.164	4.691	-12.7	102	0.00
86 T	p-Isopropyltoluene	3.384	3.875	-14.5	101	0.00
87 T	1,3-Dichlorobenzene	1.749	1.842	-5.3	97	0.00
88 T	1,4-Dichlorobenzene	1.745	1.784	-2.2	95	0.00
89 T	n-Butylbenzene	3.320	3.796	-14.3	103	0.00
90 T	Hexachloroethane	0.673	0.736	-9.4	103	0.00
91 T	1,2-Dichlorobenzene	1.497	1.532	-2.3	94	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.119	0.117	1.7	92	0.00
93 T	1,2,4-Trichlorobenzene	0.874	0.911	-4.2	96	0.00
94 T	Hexachlorobutadiene	0.477	0.526	-10.3	105	0.00
95 T	Naphthalene	1.588	1.581	0.4	84	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021122\
Data File : VD071928.D
Acq On : 11 Feb 2022 08:58
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Feb 11 23:54:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 27 06:04:38 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.749	0.785	-4.8	94	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6