

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD021420\
 Data File : VD065184.D
 Acq On : 14 Feb 2020 10:16
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 15 00:14:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D020520S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 05 15:53:00 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	75	0.00
2 T	Dichlorodifluoromethane	0.416	0.387	7.0	67	0.00
3 P	Chloromethane	0.490	0.510	-4.1	80	0.00
4 C	Vinyl Chloride	0.518	0.529	-2.1#	75	0.00
5 T	Bromomethane	0.336	0.326	3.0	74	0.00
6 T	Chloroethane	0.321	0.334	-4.0	76	0.00
7 T	Trichlorofluoromethane	0.754	0.741	1.7	71	0.00
8 T	Diethyl Ether	0.216	0.209	3.2	73	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.461	0.446	3.3	71	0.00
10 T	Methyl Iodide	0.526	0.519	1.3	70	0.00
11 T	Tert butyl alcohol	0.026	0.022	15.4	68	-0.01
12 CM	1,1-Dichloroethene	0.441	0.430	2.5#	73	0.00
13 T	Acrolein	0.032	0.024	25.0#	59	0.00
14 T	Allyl chloride	0.734	0.738	-0.5	75	0.00
15 T	Acrylonitrile	0.095	0.089	6.3	72	0.00
16 T	Acetone	0.106	0.111	-4.7	81	0.00
17 T	Carbon Disulfide	1.417	1.440	-1.6	75	0.00
18 T	Methyl Acetate	0.222	0.210	5.4	75	0.00
19 T	Methyl tert-butyl Ether	0.976	0.915	6.2	71	0.00
20 T	Methylene Chloride	0.537	0.506	5.8	78	0.00
21 T	trans-1,2-Dichloroethene	0.499	0.521	-4.4	79	0.00
22 T	Diisopropyl ether	1.384	1.476	-6.6	79	0.00
23 T	Vinyl Acetate	0.771	0.772	-0.1	73	0.00
24 P	1,1-Dichloroethane	0.837	0.891	-6.5	80	0.00
25 T	2-Butanone	0.127	0.125	1.6	77	0.00
26 T	2,2-Dichloropropane	0.751	0.781	-4.0	75	0.00
27 T	cis-1,2-Dichloroethene	0.530	0.548	-3.4	77	0.00
28 T	Bromochloromethane	0.317	0.363	-14.5	87	0.00
29 T	Tetrahydrofuran	0.077	0.071	7.8	72	0.00
30 C	Chloroform	0.834	0.894	-7.2#	80	0.00
31 T	Cyclohexane	0.845	0.775	8.3	70	0.00
32 T	1,1,1-Trichloroethane	0.777	0.801	-3.1	76	0.00
33 S	1,2-Dichloroethane-d4	0.425	0.451	-6.1	77	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	75	0.00
35 S	Dibromofluoromethane	0.279	0.329	-17.9	83	0.00
36 T	1,1-Dichloropropene	0.465	0.493	-6.0	77	0.00
37 T	Ethyl Acetate	0.190	0.177	6.8	72	0.00
38 T	Carbon Tetrachloride	0.475	0.492	-3.6	75	0.00
39 T	Methylcyclohexane	0.574	0.551	4.0	68	0.00
40 TM	Benzene	1.284	1.390	-8.3	80	0.00
41 T	Methacrylonitrile	0.105	0.118	-12.4	84	0.00
42 TM	1,2-Dichloroethane	0.366	0.366	0.0	74	0.00
43 T	Isopropyl Acetate	0.363	0.326	10.2	67	0.00
44 TM	Trichloroethene	0.369	0.382	-3.5	77	0.00
45 C	1,2-Dichloropropane	0.308	0.341	-10.7#	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD021420\
 Data File : VD065184.D
 Acq On : 14 Feb 2020 10:16
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 15 00:14:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D020520S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 05 15:53:00 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.162	0.168	-3.7	77	0.00
47 T	Bromodichloromethane	0.426	0.459	-7.7	80	0.00
48 T	Methyl methacrylate	0.175	0.158	9.7	69	0.00
49 T	1,4-Dioxane	0.002	0.002	0.0	76	0.00
50 S	Toluene-d8	1.110	1.288	-16.0	81	0.00
51 T	4-Methyl-2-Pentanone	0.179	0.174	2.8	73	0.00
52 CM	Toluene	0.819	0.894	-9.2#	79	0.00
53 T	t-1,3-Dichloropropene	0.415	0.424	-2.2	75	0.00
54 T	cis-1,3-Dichloropropene	0.496	0.527	-6.3	78	0.00
55 T	1,1,2-Trichloroethane	0.222	0.233	-5.0	79	0.00
56 T	Ethyl methacrylate	0.281	0.279	0.7	72	0.00
57 T	1,3-Dichloropropane	0.390	0.405	-3.8	78	0.00
58 T	2-Chloroethyl Vinyl ether	0.096	0.085	11.5	58	0.00
59 T	2-Hexanone	0.127	0.135	-6.3	78	0.00
60 T	Dibromochloromethane	0.287	0.303	-5.6	78	0.00
61 T	1,2-Dibromoethane	0.217	0.219	-0.9	76	0.00
62 S	4-Bromofluorobenzene	0.352	0.400	-13.6	80	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	77	0.00
64 T	Tetrachloroethene	0.359	0.356	0.8	74	0.00
65 PM	Chlorobenzene	0.967	1.031	-6.6	80	0.00
66 T	1,1,1,2-Tetrachloroethane	0.357	0.379	-6.2	79	0.00
67 C	Ethyl Benzene	1.757	1.864	-6.1#	78	0.00
68 T	m/p-Xylenes	0.671	0.725	-8.0	79	0.00
69 T	o-Xylene	0.604	0.641	-6.1	78	0.00
70 T	Styrene	1.050	1.142	-8.8	79	0.00
71 P	Bromoform	0.192	0.193	-0.5	77	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	79	0.00
73 T	Isopropylbenzene	3.432	3.530	-2.9	78	0.00
74 T	N-amyl acetate	0.746	0.676	9.4	70	0.00
75 P	1,1,2,2-Tetrachloroethane	0.540	0.526	2.6	76	0.00
76 T	1,2,3-Trichloropropane	0.379	0.366	3.4	74	0.00
77 T	Bromobenzene	0.827	0.828	-0.1	78	0.00
78 T	n-propylbenzene	4.019	4.258	-5.9	80	0.00
79 T	2-Chlorotoluene	2.240	2.305	-2.9	79	0.00
80 T	1,3,5-Trimethylbenzene	2.808	2.939	-4.7	80	0.00
81 T	trans-1,4-Dichloro-2-butene	0.189	0.176	6.9	72	0.00
82 T	4-Chlorotoluene	2.353	2.463	-4.7	81	0.00
83 T	tert-Butylbenzene	2.422	2.503	-3.3	80	0.00
84 T	1,2,4-Trimethylbenzene	2.787	2.908	-4.3	80	0.00
85 T	sec-Butylbenzene	3.308	3.453	-4.4	78	0.00
86 T	p-Isopropyltoluene	3.074	3.227	-5.0	78	0.00
87 T	1,3-Dichlorobenzene	1.548	1.617	-4.5	81	0.00
88 T	1,4-Dichlorobenzene	1.541	1.554	-0.8	79	0.00
89 T	n-Butylbenzene	2.837	2.960	-4.3	78	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD021420\
Data File : VD065184.D
Acq On : 14 Feb 2020 10:16
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Feb 15 00:14:53 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D020520S.M
Quant Title : SW846 8260
QLast Update : Wed Feb 05 15:53:00 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.582	0.614	-5.5	80	0.00
91 T	1,2-Dichlorobenzene	1.326	1.341	-1.1	79	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.088	0.082	6.8	73	0.00
93 T	1,2,4-Trichlorobenzene	0.948	0.905	4.5	74	0.00
94 T	Hexachlorobutadiene	0.583	0.563	3.4	75	0.00
95 T	Naphthalene	1.575	1.369	13.1	68	0.00
96 T	1,2,3-Trichlorobenzene	0.811	0.748	7.8	73	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6