

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021521\
 Data File : VD068360.D
 Acq On : 15 Feb 2021 13:49
 Operator : VA/SY
 Sample : VD0215SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0215SBL01

Quant Time: Feb 16 00:14:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 12 13:30:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	488568	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	893452	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	812174	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	352874	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	259869	52.03	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	104.06%
35) Dibromofluoromethane	7.920	113	263372	49.66	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.32%
50) Toluene-d8	10.344	98	1120087	50.96	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.92%
62) 4-Bromofluorobenzene	12.632	95	347796	48.57	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.14%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021521\
Data File : VD068360.D
Acq On : 15 Feb 2021 13:49
Operator : VA/SY
Sample : VD0215SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0215SBL01

Quant Time: Feb 16 00:14:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021221S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 12 13:30:51 2021
Response via : Initial Calibration

