



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_D\Data\VD021522\  
 Data File : VD071969.D  
 Acq On : 15 Feb 2022 09:21  
 Operator : VA/SY  
 Sample : BFB  
 Misc : 5.00G/5.00ml/MSVOA\_D/SOIL  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_D  
 ClientSampleId :  
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA\_D\Method\82D012622S.M  
 Title : SW846 8260  
 Last Update : Thu Jan 27 06:04:38 2022



AutoFind: Scans 1819, 1820, 1821; Background Corrected with Scan 1811

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 50             | 95              | 15              | 40              | 23.9         | 65053      | PASS                |
| 75             | 95              | 30              | 60              | 56.2         | 152704     | PASS                |
| 95             | 95              | 100             | 100             | 100.0        | 271680     | PASS                |
| 96             | 95              | 5               | 9               | 6.2          | 16721      | PASS                |
| 173            | 174             | 0.00            | 2               | 1.4          | 2698       | PASS                |
| 174            | 95              | 50              | 100             | 72.4         | 196821     | PASS                |
| 175            | 174             | 5               | 9               | 7.9          | 15584      | PASS                |
| 176            | 174             | 95              | 101             | 97.9         | 192725     | PASS                |
| 177            | 176             | 5               | 9               | 6.4          | 12330      | PASS                |