

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021621\
 Data File : VD068367.D
 Acq On : 16 Feb 2021 10:21
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 17 00:02:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 12 13:30:51 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	87	0.00
2 T	Dichlorodifluoromethane	0.458	0.426	7.0	91	0.00
3 P	Chloromethane	0.545	0.523	4.0	93	0.00
4 C	Vinyl Chloride	0.653	0.633	3.1#	92	0.00
5 T	Bromomethane	0.467	0.414	11.3	88	0.00
6 T	Chloroethane	0.427	0.404	5.4	89	0.00
7 T	Trichlorofluoromethane	0.804	0.764	5.0	89	0.00
8 T	Diethyl Ether	0.236	0.231	2.1	89	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.489	0.485	0.8	93	-0.01
10 T	Methyl Iodide	0.459	0.442	3.7	84	0.00
11 T	Tert butyl alcohol	0.026	0.027	-3.8	96	0.00
12 CM	1,1-Dichloroethene	0.476	0.457	4.0#	89	0.00
13 T	Acrolein	0.029	0.035	-20.7	96	0.00
14 T	Allyl chloride	0.618	0.635	-2.8	94	0.00
15 T	Acrylonitrile	0.102	0.106	-3.9	95	0.00
16 T	Acetone	0.079	0.090	-13.9	100	0.00
17 T	Carbon Disulfide	1.377	1.270	7.8	89	0.00
18 T	Methyl Acetate	0.216	0.218	-0.9	94	0.00
19 T	Methyl tert-butyl Ether	1.073	1.084	-1.0	92	0.00
20 T	Methylene Chloride	0.628	0.536	14.6	87	0.00
21 T	trans-1,2-Dichloroethene	0.550	0.523	4.9	89	0.00
22 T	Diisopropyl ether	1.368	1.429	-4.5	95	0.00
23 T	Vinyl Acetate	0.706	0.756	-7.1	94	0.00
24 P	1,1-Dichloroethane	0.936	0.937	-0.1	91	0.00
25 T	2-Butanone	0.110	0.119	-8.2	96	0.00
26 T	2,2-Dichloropropane	0.801	0.803	-0.2	93	0.00
27 T	cis-1,2-Dichloroethene	0.626	0.599	4.3	89	0.00
28 T	Bromochloromethane	0.388	0.416	-7.2	95	-0.01
29 T	Tetrahydrofuran	0.071	0.072	-1.4	93	0.00
30 C	Chloroform	0.974	0.963	1.1#	91	0.00
31 T	Cyclohexane	0.843	0.798	5.3	94	0.00
32 T	1,1,1-Trichloroethane	0.841	0.837	0.5	91	0.00
33 S	1,2-Dichloroethane-d4	0.511	0.569	-11.4	96	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	87	0.00
35 S	Dibromofluoromethane	0.297	0.335	-12.8	95	0.00
36 T	1,1-Dichloropropene	0.442	0.432	2.3	92	0.00
37 T	Ethyl Acetate	0.149	0.154	-3.4	92	0.00
38 T	Carbon Tetrachloride	0.422	0.411	2.6	90	0.00
39 T	Methylcyclohexane	0.543	0.533	1.8	92	0.00
40 TM	Benzene	1.262	1.240	1.7	90	0.00
41 T	Methacrylonitrile	0.086	0.101	-17.4	98	0.00
42 TM	1,2-Dichloroethane	0.337	0.343	-1.8	92	0.00
43 T	Isopropyl Acetate	0.286	0.313	-9.4	95	0.00
44 TM	Trichloroethene	0.343	0.334	2.6	89	0.00
45 C	1,2-Dichloropropane	0.308	0.319	-3.6#	95	0.00
46 T	Dibromomethane	0.167	0.165	1.2	91	0.00
47 T	Bromodichloromethane	0.433	0.438	-1.2	90	0.00
48 T	Methyl methacrylate	0.143	0.160	-11.9	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021621\
 Data File : VD068367.D
 Acq On : 16 Feb 2021 10:21
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 17 00:02:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 12 13:30:51 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	86	0.00
50 S	Toluene-d8	1.230	1.373	-11.6	95	0.00
51 T	4-Methyl-2-Pentanone	0.147	0.156	-6.1	93	0.00
52 CM	Toluene	0.811	0.795	2.0#	90	0.00
53 T	t-1,3-Dichloropropene	0.401	0.413	-3.0	92	0.00
54 T	cis-1,3-Dichloropropene	0.490	0.491	-0.2	91	0.00
55 T	1,1,2-Trichloroethane	0.231	0.237	-2.6	93	0.00
56 T	Ethyl methacrylate	0.274	0.286	-4.4	91	0.00
57 T	1,3-Dichloropropane	0.405	0.407	-0.5	92	0.00
58 T	2-Chloroethyl Vinyl ether	0.147	0.163	-10.9	98	0.00
59 T	2-Hexanone	0.099	0.112	-13.1	99	0.00
60 T	Dibromochloromethane	0.286	0.282	1.4	89	0.00
61 T	1,2-Dibromoethane	0.223	0.221	0.9	90	0.00
62 S	4-Bromofluorobenzene	0.401	0.441	-10.0	97	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	88	0.00
64 T	Tetrachloroethene	0.308	0.292	5.2	90	0.00
65 PM	Chlorobenzene	0.961	0.939	2.3	90	0.00
66 T	1,1,1,2-Tetrachloroethane	0.335	0.333	0.6	90	0.00
67 C	Ethyl Benzene	1.724	1.734	-0.6#	91	0.00
68 T	m/p-Xylenes	0.651	0.647	0.6	91	0.00
69 T	o-Xylene	0.598	0.597	0.2	90	0.00
70 T	Styrene	1.040	1.041	-0.1	91	0.00
71 P	Bromoform	0.172	0.170	1.2	89	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
73 T	Isopropylbenzene	3.652	3.699	-1.3	92	0.00
74 T	N-amyl acetate	0.645	0.692	-7.3	94	0.00
75 P	1,1,2,2-Tetrachloroethane	0.633	0.638	-0.8	92	0.00
76 T	1,2,3-Trichloropropane	0.476	0.465	2.3	91	0.00
77 T	Bromobenzene	0.816	0.796	2.5	90	0.00
78 T	n-propylbenzene	4.372	4.418	-1.1	93	0.00
79 T	2-Chlorotoluene	2.421	2.407	0.6	91	0.00
80 T	1,3,5-Trimethylbenzene	2.984	3.016	-1.1	91	0.00
81 T	trans-1,4-Dichloro-2-butene	0.190	0.196	-3.2	92	0.00
82 T	4-Chlorotoluene	2.560	2.566	-0.2	92	0.00
83 T	tert-Butylbenzene	2.589	2.592	-0.1	90	0.00
84 T	1,2,4-Trimethylbenzene	2.973	3.003	-1.0	91	0.00
85 T	sec-Butylbenzene	3.621	3.679	-1.6	90	0.00
86 T	p-Isopropyltoluene	3.251	3.357	-3.3	92	0.00
87 T	1,3-Dichlorobenzene	1.578	1.563	1.0	91	0.00
88 T	1,4-Dichlorobenzene	1.571	1.538	2.1	92	0.00
89 T	n-Butylbenzene	3.172	3.247	-2.4	93	0.00
90 T	Hexachloroethane	0.590	0.600	-1.7	92	0.00
91 T	1,2-Dichlorobenzene	1.369	1.353	1.2	91	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.093	0.093	0.0	89	0.00
93 T	1,2,4-Trichlorobenzene	0.896	0.907	-1.2	91	0.00
94 T	Hexachlorobutadiene	0.493	0.499	-1.2	91	0.00
95 T	Naphthalene	1.667	1.737	-4.2	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021621\
Data File : VD068367.D
Acq On : 16 Feb 2021 10:21
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Feb 17 00:02:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021221S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 12 13:30:51 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.782	0.773	1.2	90	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6