

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021622\
 Data File : VD071991.D
 Acq On : 16 Feb 2022 14:41
 Operator : VA/SY
 Sample : N1548-02
 Misc : 6.61G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-2(1-1.5)

Quant Time: Feb 17 00:49:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 27 06:04:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	366426	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	684605	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	665949	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	280300	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	249265	53.820	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 107.640%	
35) Dibromofluoromethane	7.914	113	224750	49.943	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 99.880%	
50) Toluene-d8	10.338	98	900459	53.578	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 107.160%	
62) 4-Bromofluorobenzene	12.626	95	306271	52.184	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 104.360%	
Target Compounds						
64) Tetrachloroethene	10.873	164	112506	23.884	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021622\
Data File : VD071991.D
Acq On : 16 Feb 2022 14:41
Operator : VA/SY
Sample : N1548-02
Misc : 6.61G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-2(1-1.5)

Quant Time: Feb 17 00:49:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 27 06:04:38 2022
Response via : Initial Calibration

