

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021622\
 Data File : VD071999.D
 Acq On : 16 Feb 2022 18:47
 Operator : VA/SY
 Sample : N1564-04
 Misc : 6.30G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-3

Quant Time: Feb 17 00:51:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 27 06:04:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	324201	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	612632	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	595554	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	256451	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	220805	53.885	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	107.760%
35) Dibromofluoromethane	7.909	113	196707	48.847	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.700%
50) Toluene-d8	10.338	98	779712	51.844	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.680%
62) 4-Bromofluorobenzene	12.620	95	277697	52.875	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	105.740%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021622\
Data File : VD071999.D
Acq On : 16 Feb 2022 18:47
Operator : VA/SY
Sample : N1564-04
Misc : 6.30G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-3

Quant Time: Feb 17 00:51:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D012622S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 27 06:04:38 2022
Response via : Initial Calibration

