

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD021720\
 Data File : VD065209.D
 Acq On : 17 Feb 2020 21:03
 Operator : VA/SY
 Sample : L1545-02RE
 Misc : 4.99G/5.00ml/MSVOA D/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 43ST-ARE

Quant Time: Feb 18 04:09:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D020520S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 05 15:53:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	282246	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	483009	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	410313	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	131640	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.34	65	171875	71.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	143.30%	
35) Dibromofluoromethane	7.92	113	107672	39.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.90%	
50) Toluene-d8	10.34	98	532376	49.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.26%	
62) 4-Bromofluorobenzene	12.64	95	138113	40.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.28%	
Target Compounds						
16) Acetone	4.17	43	9565	15.948	ug/l	99
20) Methylene Chloride	4.97	84	64039	22.047	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD021720\
Data File : VD065209.D
Acq On : 17 Feb 2020 21:03
Operator : VA/SY
Sample : L1545-02RE
Misc : 4.99G/5.00ml/MSVOA D/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
43ST-ARE

Quant Time: Feb 18 04:09:01 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D020520S.M
Quant Title : SW846 8260
QLast Update : Wed Feb 05 15:53:00 2020
Response via : Initial Calibration

