

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021721\
 Data File : VD068377.D
 Acq On : 17 Feb 2021 12:29
 Operator : VA/SY
 Sample : VD0217SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0217SBL01

Quant Time: Feb 18 00:10:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 12 13:30:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	449608	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	845939	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	770909	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	343253	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	250580	54.52	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	109.04%
35) Dibromofluoromethane	7.920	113	252930	50.37	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.74%
50) Toluene-d8	10.344	98	1071739	51.50	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.00%
62) 4-Bromofluorobenzene	12.632	95	331761	48.94	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.88%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021721\
Data File : VD068377.D
Acq On : 17 Feb 2021 12:29
Operator : VA/SY
Sample : VD0217SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0217SBL01

Quant Time: Feb 18 00:10:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021221S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 12 13:30:51 2021
Response via : Initial Calibration

