

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021721\
 Data File : VD068385.D
 Acq On : 17 Feb 2021 16:14
 Operator : VA/SY
 Sample : M1440-01
 Misc : 6.98G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SP-227-233-A

Quant Time: Feb 18 00:13:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 12 13:30:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	388372	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	754360	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	707764	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	294297	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	240647	60.61	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 121.22%	
35) Dibromofluoromethane	7.920	113	232294	51.88	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 103.76%	
50) Toluene-d8	10.344	98	961878	51.83	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 103.66%	
62) 4-Bromofluorobenzene	12.632	95	302634	50.06	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 100.12%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021721\
Data File : VD068385.D
Acq On : 17 Feb 2021 16:14
Operator : VA/SY
Sample : M1440-01
Misc : 6.98G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SP-227-233-A

Quant Time: Feb 18 00:13:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021221S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 12 13:30:51 2021
Response via : Initial Calibration

