

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021721\
 Data File : VD068387.D
 Acq On : 17 Feb 2021 17:08
 Operator : VA/SY
 Sample : M1441-01
 Misc : 6.16G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SP-234-A

Quant Time: Feb 18 00:13:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 12 13:30:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	382454	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	748046	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	701983	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	306168	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	233249	59.66	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.32%
35) Dibromofluoromethane	7.920	113	232995	52.47	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.94%
50) Toluene-d8	10.344	98	952702	51.77	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.54%
62) 4-Bromofluorobenzene	12.632	95	299237	49.92	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	99.84%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021721\
Data File : VD068387.D
Acq On : 17 Feb 2021 17:08
Operator : VA/SY
Sample : M1441-01
Misc : 6.16G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SP-234-A

Quant Time: Feb 18 00:13:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021221S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 12 13:30:51 2021
Response via : Initial Calibration

