


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD021820\
Data File : VD065211.D
Acq On : 18 Feb 2020 09:31
Operator : VA/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D021820S.M
Title : SW846 8260
Last Update : Wed Feb 19 03:26:54 2020



AutoFind: Scans 1821, 1822, 1823; Background Corrected with Scan 1813

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.9	43584	PASS
75	95	30	60	50.8	117040	PASS
95	95	100	100	100.0	230357	PASS
96	95	5	9	6.6	15178	PASS
173	174	0.00	2	1.5	2997	PASS
174	95	50	100	89.2	205440	PASS
175	174	5	9	7.8	16087	PASS
176	174	95	101	95.8	196757	PASS
177	176	5	9	7.0	13679	PASS