



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021822\
 Data File : VD072015.D
 Acq On : 18 Feb 2022 08:52
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
 Title : SW846 8260
 Last Update : Sat Feb 19 04:05:38 2022



AutoFind: Scans 1819, 1820, 1821; Background Corrected with Scan 1810

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.0	79688	PASS
75	95	30	60	50.3	210645	PASS
95	95	100	100	100.0	418560	PASS
96	95	5	9	6.5	27368	PASS
173	174	0.00	2	1.3	4243	PASS
174	95	50	100	77.6	324629	PASS
175	174	5	9	7.8	25389	PASS
176	174	95	101	97.9	317888	PASS
177	176	5	9	6.4	20192	PASS