

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021921\
 Data File : VD068403.D
 Acq On : 19 Feb 2021 16:33
 Operator : VA/SY
 Sample : M1452-13
 Misc : 6.06G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-4

Quant Time: Feb 19 23:31:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 12 13:30:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	385112	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	734961	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	677004	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	281574	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	227691	57.83	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.66%
35) Dibromofluoromethane	7.920	113	225407	51.67	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.34%
50) Toluene-d8	10.343	98	941809	52.09	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.18%
62) 4-Bromofluorobenzene	12.631	95	292115	49.60	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	99.20%
Target Compounds						
16) Acetone	4.155	43	8640	14.22	ug/l	Qvalue # 74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021921\
Data File : VD068403.D
Acq On : 19 Feb 2021 16:33
Operator : VA/SY
Sample : M1452-13
Misc : 6.06G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-4

Quant Time: Feb 19 23:31:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021221S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 12 13:30:51 2021
Response via : Initial Calibration

