

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021921\
 Data File : VD068404.D
 Acq On : 19 Feb 2021 17:00
 Operator : VA/SY
 Sample : M1452-17
 Misc : 6.33G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-5

Quant Time: Feb 19 23:32:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 12 13:30:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	378491	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	728167	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	672824	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	291665	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	225789	58.35	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 116.70%	
35) Dibromofluoromethane	7.920	113	217591	50.34	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 100.68%	
50) Toluene-d8	10.338	98	916988	51.19	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 102.38%	
62) 4-Bromofluorobenzene	12.632	95	289907	49.68	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 99.36%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021921\
Data File : VD068404.D
Acq On : 19 Feb 2021 17:00
Operator : VA/SY
Sample : M1452-17
Misc : 6.33G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-5

Quant Time: Feb 19 23:32:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021221S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 12 13:30:51 2021
Response via : Initial Calibration

