

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021921\
 Data File : VD068394.D
 Acq On : 19 Feb 2021 11:54
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 19 23:28:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 12 13:30:51 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	76	0.00
2 T	Dichlorodifluoromethane	0.458	0.386	15.7	72	0.00
3 P	Chloromethane	0.545	0.483	11.4	75	0.00
4 C	Vinyl Chloride	0.653	0.576	11.8#	73	0.00
5 T	Bromomethane	0.467	0.397	15.0	74	0.00
6 T	Chloroethane	0.427	0.380	11.0	73	0.00
7 T	Trichlorofluoromethane	0.804	0.714	11.2	73	0.00
8 T	Diethyl Ether	0.236	0.219	7.2	74	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.489	0.433	11.5	73	-0.01
10 T	Methyl Iodide	0.459	0.393	14.4	66	0.00
11 T	Tert butyl alcohol	0.026	0.034	-30.8#	107	-0.01
12 CM	1,1-Dichloroethene	0.476	0.408	14.3#	70	0.00
13 T	Acrolein	0.029	0.030	-3.4	73	0.00
14 T	Allyl chloride	0.618	0.584	5.5	76	0.00
15 T	Acrylonitrile	0.102	0.098	3.9	77	0.00
16 T	Acetone	0.079	0.077	2.5	75	0.00
17 T	Carbon Disulfide	1.377	1.172	14.9	72	0.00
18 T	Methyl Acetate	0.216	0.210	2.8	79	0.00
19 T	Methyl tert-butyl Ether	1.073	0.999	6.9	74	0.00
20 T	Methylene Chloride	0.628	0.511	18.6	72	0.00
21 T	trans-1,2-Dichloroethene	0.550	0.496	9.8	74	0.00
22 T	Diisopropyl ether	1.368	1.336	2.3	78	0.00
23 T	Vinyl Acetate	0.706	0.695	1.6	76	0.00
24 P	1,1-Dichloroethane	0.936	0.872	6.8	74	0.00
25 T	2-Butanone	0.110	0.108	1.8	76	0.00
26 T	2,2-Dichloropropane	0.801	0.760	5.1	77	0.00
27 T	cis-1,2-Dichloroethene	0.626	0.581	7.2	76	0.00
28 T	Bromochloromethane	0.388	0.427	-10.1	85	0.00
29 T	Tetrahydrofuran	0.071	0.068	4.2	77	0.00
30 C	Chloroform	0.974	0.917	5.9#	76	0.00
31 T	Cyclohexane	0.843	0.722	14.4	75	0.00
32 T	1,1,1-Trichloroethane	0.841	0.777	7.6	74	0.00
33 S	1,2-Dichloroethane-d4	0.511	0.531	-3.9	78	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	77	0.00
35 S	Dibromofluoromethane	0.297	0.317	-6.7	80	0.00
36 T	1,1-Dichloropropene	0.442	0.401	9.3	76	0.00
37 T	Ethyl Acetate	0.149	0.147	1.3	78	0.00
38 T	Carbon Tetrachloride	0.422	0.389	7.8	75	0.00
39 T	Methylcyclohexane	0.543	0.484	10.9	74	0.00
40 TM	Benzene	1.262	1.164	7.8	75	0.00
41 T	Methacrylonitrile	0.086	0.096	-11.6	83	0.01
42 TM	1,2-Dichloroethane	0.337	0.313	7.1	75	0.00
43 T	Isopropyl Acetate	0.286	0.275	3.8	75	0.00
44 TM	Trichloroethene	0.343	0.315	8.2	74	0.00
45 C	1,2-Dichloropropane	0.308	0.295	4.2#	78	0.00
46 T	Dibromomethane	0.167	0.155	7.2	76	0.00
47 T	Bromodichloromethane	0.433	0.416	3.9	76	0.00
48 T	Methyl methacrylate	0.143	0.128	10.5	76	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021921\
 Data File : VD068394.D
 Acq On : 19 Feb 2021 11:54
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 19 23:28:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 12 13:30:51 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	70	0.00
50 S	Toluene-d8	1.230	1.265	-2.8	78	0.00
51 T	4-Methyl-2-Pentanone	0.147	0.146	0.7	77	0.00
52 CM	Toluene	0.811	0.754	7.0#	76	0.00
53 T	t-1,3-Dichloropropene	0.401	0.397	1.0	78	0.00
54 T	cis-1,3-Dichloropropene	0.490	0.459	6.3	76	0.00
55 T	1,1,2-Trichloroethane	0.231	0.223	3.5	78	0.00
56 T	Ethyl methacrylate	0.274	0.261	4.7	74	0.00
57 T	1,3-Dichloropropane	0.405	0.381	5.9	77	0.00
58 T	2-Chloroethyl Vinyl ether	0.147	0.152	-3.4	81	0.00
59 T	2-Hexanone	0.099	0.097	2.0	76	0.00
60 T	Dibromochloromethane	0.286	0.272	4.9	77	0.00
61 T	1,2-Dibromoethane	0.223	0.203	9.0	73	0.00
62 S	4-Bromofluorobenzene	0.401	0.409	-2.0	80	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	78	0.00
64 T	Tetrachloroethene	0.308	0.273	11.4	74	0.00
65 PM	Chlorobenzene	0.961	0.896	6.8	76	0.00
66 T	1,1,1,2-Tetrachloroethane	0.335	0.321	4.2	77	0.00
67 C	Ethyl Benzene	1.724	1.628	5.6#	76	0.00
68 T	m/p-Xylenes	0.651	0.610	6.3	76	0.00
69 T	o-Xylene	0.598	0.564	5.7	75	0.00
70 T	Styrene	1.040	0.977	6.1	76	0.00
71 P	Bromoform	0.172	0.161	6.4	74	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	80	0.00
73 T	Isopropylbenzene	3.652	3.379	7.5	75	0.00
74 T	N-amyl acetate	0.645	0.643	0.3	78	0.00
75 P	1,1,2,2-Tetrachloroethane	0.633	0.591	6.6	77	0.00
76 T	1,2,3-Trichloropropane	0.476	0.427	10.3	75	0.00
77 T	Bromobenzene	0.816	0.742	9.1	75	0.00
78 T	n-propylbenzene	4.372	4.071	6.9	77	0.00
79 T	2-Chlorotoluene	2.421	2.242	7.4	76	0.00
80 T	1,3,5-Trimethylbenzene	2.984	2.784	6.7	75	0.00
81 T	trans-1,4-Dichloro-2-butene	0.190	0.184	3.2	77	0.00
82 T	4-Chlorotoluene	2.560	2.346	8.4	76	0.00
83 T	tert-Butylbenzene	2.589	2.441	5.7	76	0.00
84 T	1,2,4-Trimethylbenzene	2.973	2.813	5.4	77	0.00
85 T	sec-Butylbenzene	3.621	3.426	5.4	76	0.00
86 T	p-Isopropyltoluene	3.251	3.100	4.6	77	0.00
87 T	1,3-Dichlorobenzene	1.578	1.468	7.0	77	0.00
88 T	1,4-Dichlorobenzene	1.571	1.453	7.5	78	0.00
89 T	n-Butylbenzene	3.172	2.982	6.0	77	0.00
90 T	Hexachloroethane	0.590	0.573	2.9	79	0.00
91 T	1,2-Dichlorobenzene	1.369	1.271	7.2	76	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.093	0.087	6.5	74	0.00
93 T	1,2,4-Trichlorobenzene	0.896	0.847	5.5	76	0.00
94 T	Hexachlorobutadiene	0.493	0.474	3.9	78	0.00
95 T	Naphthalene	1.667	1.574	5.6	75	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD021921\
Data File : VD068394.D
Acq On : 19 Feb 2021 11:54
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Feb 19 23:28:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021221S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 12 13:30:51 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.782	0.729	6.8	76	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6