

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022123\
 Data File : VD075382.D
 Acq On : 21 Feb 2023 15:55
 Operator : KP/SY
 Sample : IBLK
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 IBLK

Quant Time: Feb 21 17:57:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022123S.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 21 16:26:04 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	58981	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	100945	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.586	117	89695	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	39635	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	26352	52.018	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 104.040%	
35) Dibromofluoromethane	7.804	113	27443	45.795	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 91.600%	
50) Toluene-d8	10.269	98	90509	41.851	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 83.700%	
62) 4-Bromofluorobenzene	12.580	95	30153	41.522	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 83.040%	
Target Compounds						
					Qvalue	
16) Acetone	4.022	43	192	1.258	ug/l #	44
20) Methylene Chloride	4.787	84	1582	Below Cal	#	87
31) Cyclohexane	7.881	56	1962	1.639	ug/l #	65
36) 1,1-Dichloropropene	7.869	75	5348	4.979	ug/l #	48
38) Carbon Tetrachloride	7.875	117	5918	5.499	ug/l #	19
76) 1,2,3-Trichloropropane	12.574	75	16371	55.921	ug/l #	100
81) trans-1,4-Dichloro-2-b...	12.574	75	16371	125.126	ug/l #	5

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022123\
Data File : VD075382.D
Acq On : 21 Feb 2023 15:55
Operator : KP/SY
Sample : IBLK
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
IBLK

Quant Time: Feb 21 17:57:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022123S.M
Quant Title : SW846 8260
QLast Update : Tue Feb 21 16:26:04 2023
Response via : Initial Calibration

