

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022123\
 Data File : VD075384.D
 Acq On : 21 Feb 2023 16:50
 Operator : KP/SY
 Sample : 01613-06
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BP-VPB-RW10-GW-838-840

Quant Time: Feb 21 17:59:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022123S.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 21 16:26:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	47384	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	80197	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.587	117	68642	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	26146	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	20342	49.849	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.700%
35) Dibromofluoromethane	7.810	113	22018	46.248	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	92.500%
50) Toluene-d8	10.269	98	70737	41.170	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	82.340%
62) 4-Bromofluorobenzene	12.575	95	21586	37.415	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	74.840%
Target Compounds						
20) Methylene Chloride	4.805	84	6799	6.079	ug/l	# 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022123\
Data File : VD075384.D
Acq On : 21 Feb 2023 16:50
Operator : KP/SY
Sample : 01613-06
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BP-VPB-RW10-GW-838-840

Quant Time: Feb 21 17:59:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022123S.M
Quant Title : SW846 8260
QLast Update : Tue Feb 21 16:26:04 2023
Response via : Initial Calibration

