

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022123\
 Data File : VD075385.D
 Acq On : 21 Feb 2023 17:17
 Operator : KP/SY
 Sample : 01613-07
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BP-VPB-RW10-GW-908-910

Quant Time: Feb 21 17:59:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022123S.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 21 16:26:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	18859	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	34165	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.587	117	33395	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	14672	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	13695	86.667	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	173.340%	#
35) Dibromofluoromethane	7.810	113	12128	59.797	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	119.600%	
50) Toluene-d8	10.275	98	33197	45.354	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	90.700%	
62) 4-Bromofluorobenzene	12.575	95	12093	49.202	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	98.400%	
Target Compounds						
					Qvalue	
16) Acetone	4.028	43	1878	38.477	ug/l #	88
20) Methylene Chloride	4.816	84	2261	3.914	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022123\
Data File : VD075385.D
Acq On : 21 Feb 2023 17:17
Operator : KP/SY
Sample : 01613-07
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BP-VPB-RW10-GW-908-910

Quant Time: Feb 21 17:59:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022123S.M
Quant Title : SW846 8260
QLast Update : Tue Feb 21 16:26:04 2023
Response via : Initial Calibration

