

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022224\
 Data File : VD078523.D
 Acq On : 22 Feb 2024 17:54
 Operator : RP/MD
 Sample : P1538-07MS
 Misc : 5.03G/5ml/MSVOA_D/SOIL/B
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 20241101-COMPOSITE-35N-N-4-06MS

Quant Time: Feb 23 01:06:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021324S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 14 01:16:20 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 02/23/2024
 Supervised By :Semsettin Yesilyurt 02/23/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	57724	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.775	114	103917	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	82257	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	28038	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	43575	69.816	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 139.640%			
35) Dibromofluoromethane	7.810	113	43347	65.667	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 131.340%			
50) Toluene-d8	10.269	98	141694	53.660	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 107.320%			
62) 4-Bromofluorobenzene	12.569	95	32349	42.098	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery = 84.200%			
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.934	85	33405	58.734	ug/l	98
3) Chloromethane	2.152	50	66162	63.300	ug/l	98
4) Vinyl Chloride	2.281	62	97497	61.129	ug/l	96
5) Bromomethane	2.687	94	47333	42.994	ug/l	99
6) Chloroethane	2.828	64	70512	60.628	ug/l	99
7) Trichlorofluoromethane	3.164	101	56252	52.283	ug/l	91
8) Diethyl Ether	3.587	74	19968	64.435	ug/l	93
9) 1,1,2-Trichlorotrifluo...	3.969	101	30257	44.578	ug/l	97
10) Methyl Iodide	4.163	142	16872	27.754	ug/l #	88
11) Tert butyl alcohol	5.058	59	13273	395.663	ug/l	97
12) 1,1-Dichloroethene	3.934	96	31196	49.575	ug/l	96
13) Acrolein	3.793	56	167	2.896	ug/l #	1
14) Allyl chloride	4.569	41	38682	49.759	ug/l	93
15) Acrylonitrile	5.252	53	48181	363.188	ug/l	99
16) Acetone	4.028	43	51738	465.475	ug/l	87
17) Carbon Disulfide	4.269	76	77541	36.620	ug/l	96
18) Methyl Acetate	4.563	43	59237	120.633	ug/l	94
19) Methyl tert-butyl Ether	5.322	73	90477	67.674	ug/l	99
20) Methylene Chloride	4.793	84	60708	88.226	ug/l	94
21) trans-1,2-Dichloroethene	5.305	96	34880	47.118	ug/l	87
22) Diisopropyl ether	6.216	45	98841	57.687	ug/l	94
23) Vinyl Acetate	6.157	43	66574	85.584	ug/l	96
24) 1,1-Dichloroethane	6.116	63	67870	55.916	ug/l	96
25) 2-Butanone	7.081	43	63336	410.610	ug/l #	85
26) 2,2-Dichloropropane	7.075	77	54050	53.120	ug/l	95
27) cis-1,2-Dichloroethene	7.081	96	42011	51.725	ug/l	92
28) Bromochloromethane	7.428	49	30214	62.848	ug/l	86
29) Tetrahydrofuran	7.440	42	35935	387.150	ug/l	90
30) Chloroform	7.593	83	75525	58.810	ug/l	98
31) Cyclohexane	7.875	56	30639	29.020	ug/l	93
32) 1,1,1-Trichloroethane	7.793	97	57189	52.231	ug/l	98
36) 1,1-Dichloropropene	8.010	75	43081	42.012	ug/l	95
37) Ethyl Acetate	7.169	43	20389	56.720	ug/l	98
38) Carbon Tetrachloride	7.987	117	25702	25.309	ug/l	98
39) Methylcyclohexane	9.269	83	22065	18.125	ug/l #	86
40) Benzene	8.251	78	156985	52.964	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022224\
 Data File : VD078523.D
 Acq On : 22 Feb 2024 17:54
 Operator : RP/MD
 Sample : P1538-07MS
 Misc : 5.03G/5ml/MSVOA_D/SOIL/B
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 20241101-COMPOSITE-35N-N-4-06MS

Quant Time: Feb 23 01:06:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021324S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 14 01:16:20 2024
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 02/23/2024
 Supervised By :Semsettin Yesilyurt 02/23/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.399	41	11162	59.665	ug/l #	80
42) 1,2-Dichloroethane	8.322	62	46839	61.073	ug/l	96
43) Isopropyl Acetate	8.363	43	43613	65.520	ug/l #	79
44) Trichloroethene	9.028	130	36678	48.505	ug/l	96
45) 1,2-Dichloropropane	9.304	63	40787	56.085	ug/l	100
46) Dibromomethane	9.387	93	25461	62.245	ug/l	91
47) Bromodichloromethane	9.587	83	56650	55.915	ug/l	98
48) Methyl methacrylate	9.381	41	27207	77.041	ug/l #	80
49) 1,4-Dioxane	9.387	88	7322	1652.894	ug/l #	85
51) 4-Methyl-2-Pentanone	10.157	43	131044	370.925	ug/l	91
52) Toluene	10.334	92	89622	47.814	ug/l	95
53) t-1,3-Dichloropropene	10.551	75	40885	44.226	ug/l	100
54) cis-1,3-Dichloropropene	10.016	75	43890	39.153	ug/l #	92
55) 1,1,2-Trichloroethane	10.734	97	35756	64.910	ug/l	91
56) Ethyl methacrylate	10.598	69	27626	54.225	ug/l #	89
57) 1,3-Dichloropropane	10.881	76	58155	62.062	ug/l	98
58) 2-Chloroethyl Vinyl ether	9.869	63	112533	374.990	ug/l	96
59) 2-Hexanone	10.922	43	94216	354.861	ug/l	90
60) Dibromochloromethane	11.069	129	37814	57.054	ug/l	94
61) 1,2-Dibromoethane	11.181	107	31658	61.216	ug/l	98
64) Tetrachloroethene	10.810	164	31228	60.105	ug/l	97
65) Chlorobenzene	11.604	112	85211	51.528	ug/l	94
66) 1,1,1,2-Tetrachloroethane	11.681	131	36504	60.067	ug/l	94
67) Ethyl Benzene	11.681	91	128922	43.511	ug/l	99
68) m/p-Xylenes	11.792	106	99437	86.036	ug/l	91
69) o-Xylene	12.122	106	48278	45.774	ug/l	92
70) Styrene	12.134	104	70501	40.957	ug/l	95
71) Bromoform	12.298	173	20551	64.248	ug/l #	99
73) Isopropylbenzene	12.422	105	95848	50.909	ug/l	96
74) N-amyl acetate	12.234	43	18568	52.131	ug/l	97
75) 1,1,2,2-Tetrachloroethane	12.675	83	40382	108.127	ug/l	98
76) 1,2,3-Trichloropropane	12.722	75	25566m	107.405	ug/l	
77) Bromobenzene	12.698	156	29914	67.607	ug/l	83
78) n-propylbenzene	12.763	91	103412	44.076	ug/l	95
79) 2-Chlorotoluene	12.845	91	66264	52.731	ug/l	95
80) 1,3,5-Trimethylbenzene	12.904	105	68942	44.099	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.469	75	4726	41.218	ug/l	95
82) 4-Chlorotoluene	12.945	91	70216	52.025	ug/l	96
83) tert-Butylbenzene	13.169	119	52552	39.859	ug/l	97
84) 1,2,4-Trimethylbenzene	13.210	105	71900	45.316	ug/l	99
85) sec-Butylbenzene	13.345	105	66726	32.633	ug/l	93
86) p-Isopropyltoluene	13.457	119	56411	32.391	ug/l	97
87) 1,3-Dichlorobenzene	13.457	146	43245	45.793	ug/l	98
88) 1,4-Dichlorobenzene	13.539	146	42553	45.860	ug/l	98
89) n-Butylbenzene	13.786	91	44928	26.764	ug/l	92
90) Hexachloroethane	14.051	117	4426	12.365	ug/l	99
91) 1,2-Dichlorobenzene	13.834	146	39347	49.617	ug/l	97
92) 1,2-Dibromo-3-Chloropr...	14.445	75	5136	105.027	ug/l	92
93) 1,2,4-Trichlorobenzene	15.098	180	11262	24.165	ug/l	99
94) Hexachlorobutadiene	15.204	225	4029	16.562	ug/l	96
95) Naphthalene	15.333	128	37268	43.113	ug/l	98
96) 1,2,3-Trichlorobenzene	15.522	180	10439	26.429	ug/l	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022224\
Data File : VD078523.D
Acq On : 22 Feb 2024 17:54
Operator : RP/MD
Sample : P1538-07MS
Misc : 5.03G/5ml/MSVOA_D/SOIL/B
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
20241101-COMPOSITE-35N-N-4-06MS


Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024

Quant Time: Feb 23 01:06:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021324S.M
Quant Title : SW846 8260
QLast Update : Wed Feb 14 01:16:20 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed







#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.875 min Scan# 1069
Delta R.T. -0.006 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54

Instrument : MSVOA_D
ClientSampleId : 20241101-COMPOSITE-35N-N-4-06MS

Tgt Ion:168 Resp: 57724
Ion Ratio Lower Upper
168 100
99 57.7 36.7 55.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024



#2
Dichlorodifluoromethane
Concen: 58.734 ug/l
RT: 1.934 min Scan# 59
Delta R.T. -0.006 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54


Tgt Ion: 85 Resp: 33405
Ion Ratio Lower Upper
85 100
87 30.3 15.7 47.0





#3
Chloromethane
Concen: 63.300 ug/l
RT: 2.152 min Scan# 90
Delta R.T. -0.006 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54




Instrument : MSVOA_D
ClientSampleId : 20241101-COMPOSITE-35N-N-4-06MS



Tgt Ion: 50 Resp: 66162
Ion Ratio Lower Upper
50 100
52 32.1 26.6 40.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024



#4
Vinyl Chloride
Concen: 61.129 ug/l
RT: 2.281 min Scan# 118
Delta R.T. -0.006 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54


Tgt Ion: 62 Resp: 97497
Ion Ratio Lower Upper
62 100
64 30.5 26.0 39.0





#5
Bromomethane
Concen: 42.994 ug/l
RT: 2.687 min Scan# 11
Delta R.T. -0.006 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54




Instrument :
MSVOA_D
ClientSampleId :
20241101-COMPOSITE-35N-N-4-06MS



Tgt Ion: 94 Resp: 4733
Ion Ratio Lower Upper
94 100
96 96.2 75.8 113.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024



#6
Chloroethane
Concen: 60.628 ug/l
RT: 2.828 min Scan# 211
Delta R.T. -0.012 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54


Tgt Ion: 64 Resp: 70512
Ion Ratio Lower Upper
64 100
66 32.6 25.5 38.3





#7
Trichlorofluoromethane
Concen: 52.283 ug/l
RT: 3.164 min Scan# 20
Delta R.T. -0.012 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54




Instrument :
MSVOA_D
ClientSampleId :
20241101-COMPOSITE-35N-N-4-06MS



Tgt Ion:101 Resp: 5625
Ion Ratio Lower Upper
101 100
103 58.0 52.3 78.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024



#8
Diethyl Ether
Concen: 64.435 ug/l
RT: 3.587 min Scan# 340
Delta R.T. -0.012 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54

Tgt Ion: 74 Resp: 19968
Ion Ratio Lower Upper
74 100
45 81.1 37.5 112.7









#9
1,1,2-Trichlorotrifluoroethane
Concen: 44.578 ug/l
RT: 3.969 min Scan# 405
Delta R.T. -0.000 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54

Instrument :
MSVOA_D
ClientSampleId :
20241101-COMPOSITE-35N-N-4-06MS

Tgt Ion:101 Resp: 3025
Ion Ratio Lower Upper
101 100
85 43.4 33.8 50.6
151 71.8 59.8 89.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024



#10
Methyl Iodide
Concen: 27.754 ug/l
RT: 4.163 min Scan# 438
Delta R.T. -0.006 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54


Tgt Ion:142 Resp: 16872
Ion Ratio Lower Upper
142 100
127 45.8 29.4 44.0#
141 14.8 11.1 16.7





#11
Tert butyl alcohol
Concen: 395.663 ug/l
RT: 5.058 min Scan# 590
Delta R.T. 0.006 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54




Instrument : MSVOA_D
ClientSampleId : 20241101-COMPOSITE-35N-N-4-06MS




Tgt Ion: 59 Resp: 1327
Ion Ratio Lower Upper
59 100
57 8.7 7.8 11.8

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024




#12
1,1-Dichloroethene
Concen: 49.575 ug/l
RT: 3.934 min Scan# 399
Delta R.T. -0.018 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54



Tgt Ion: 96 Resp: 31196
Ion Ratio Lower Upper
96 100
61 129.3 100.6 150.8
98 57.7 50.2 75.2





#13

Acrolein

Concen: 2.896 ug/l

RT: 3.793 min Scan# 31

Delta R.T. -0.006 min

Lab File: VD078523.D

Acq: 22 Feb 2024 17:54

Instrument :

MSVOA_D

ClientSampleId :

20241101-COMPOSITE-35N-N-4-06MS

Tgt Ion: 56 Resp: 16

Ion Ratio Lower Upper

56 100





55 203.0 55.1 82.7

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 02/23/2024

Supervised By :Semsettin Yesilyurt 02/23/2024



#14

Allyl chloride

Concen: 49.759 ug/l

RT: 4.569 min Scan# 507

Delta R.T. -0.000 min

Lab File: VD078523.D

Acq: 22 Feb 2024 17:54




Tgt Ion: 41 Resp: 38682


Ion Ratio Lower Upper

41 100

39 67.1 53.8 80.6


76 46.1 45.8 68.8





#15
Acrylonitrile
Concen: 363.188 ug/l
RT: 5.252 min Scan# 61
Delta R.T. -0.006 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54




Instrument :
MSVOA_D
ClientSampleId :
20241101-COMPOSITE-35N-N-4-06MS



Tgt Ion: 53 Resp: 4818
Ion Ratio Lower Upper
53 100
52 81.0 65.0 97.4
51 36.5 28.3 42.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024



#16
Acetone
Concen: 465.475 ug/l
RT: 4.028 min Scan# 415
Delta R.T. -0.006 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54


Tgt Ion: 43 Resp: 51738
Ion Ratio Lower Upper
43 100
58 35.0 34.7 52.1





#17
Carbon Disulfide
Concen: 36.620 ug/l
RT: 4.269 min Scan# 41
Delta R.T. -0.006 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54




Instrument :
MSVOA_D
ClientSampleId :
20241101-COMPOSITE-35N-N-4-06MS



Tgt Ion: 76 Resp: 7754
Ion Ratio Lower Upper
76 100
78 9.9 6.8 10.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024



#18
Methyl Acetate
Concen: 120.633 ug/l
RT: 4.563 min Scan# 506
Delta R.T. -0.012 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54


Tgt Ion: 43 Resp: 59237
Ion Ratio Lower Upper
43 100
74 29.6 26.4 39.6





#19
Methyl tert-butyl Ether
Concen: 67.674 ug/l
RT: 5.322 min Scan# 61
Delta R.T. 0.000 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54




Instrument :
MSVOA_D
ClientSampleId :
20241101-COMPOSITE-35N-N-4-06MS




Tgt Ion: 73 Resp: 9047
Ion Ratio Lower Upper
73 100
57 21.3 16.6 25.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024




#20
Methylene Chloride
Concen: 88.226 ug/l
RT: 4.793 min Scan# 545
Delta R.T. -0.012 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54




Tgt Ion: 84 Resp: 60708
Ion Ratio Lower Upper
84 100
49 97.1 71.4 107.2
51 31.7 24.5 36.7
86 58.4 48.7 73.1





#21
trans-1,2-Dichloroethene
Concen: 47.118 ug/l
RT: 5.305 min Scan# 611
Delta R.T. -0.012 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54




Instrument : MSVOA_D
ClientSampleId : 20241101-COMPOSITE-35N-N-4-06MS



Tgt Ion: 96 Resp: 34880
Ion Ratio Lower Upper
96 100
61 131.8 89.3 133.9
98 62.3 48.6 73.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024



#22
Diisopropyl ether
Concen: 57.687 ug/l
RT: 6.216 min Scan# 787
Delta R.T. -0.006 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54


Tgt Ion: 45 Resp: 98841
Ion Ratio Lower Upper
45 100
43 56.8 40.0 60.0
87 33.5 28.2 42.4
59 13.3 11.0 16.4





#23
 Vinyl Acetate
 Concen: 85.584 ug/l
 RT: 6.157 min Scan# 71
 Delta R.T. -0.006 min
 Lab File: VD078523.D
 Acq: 22 Feb 2024 17:54




Instrument :
 MSVOA_D
 ClientSampleId :
 20241101-COMPOSITE-35N-N-4-06MS



Tgt Ion: 43 Resp: 66574
 Ion Ratio Lower Upper
 43 100
 86 13.9 12.5 18.7




Manual Integrations
 APPROVED


Reviewed By :Mahesh Dadoda 02/23/2024
 Supervised By :Semsettin Yesilyurt 02/23/2024



#24
 1,1-Dichloroethane
 Concen: 55.916 ug/l
 RT: 6.116 min Scan# 770
 Delta R.T. -0.000 min
 Lab File: VD078523.D
 Acq: 22 Feb 2024 17:54

Tgt Ion: 63 Resp: 67870
 Ion Ratio Lower Upper
 63 100
 98 7.0 4.5 13.7
 100 4.4 2.5 7.5





#25

2-Butanone

Concen: 410.610 ug/l

RT: 7.081 min Scan# 91

Delta R.T. -0.006 min

Lab File: VD078523.D

Acq: 22 Feb 2024 17:54

Instrument :

MSVOA_D

ClientSampleId :

20241101-COMPOSITE-35N-N-4-06MS

Tgt Ion: 43 Resp: 63330

Ion Ratio Lower Upper

43 100





72 29.4 30.7 46.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 02/23/2024

Supervised By :Semsettin Yesilyurt 02/23/2024



#26

2,2-Dichloropropane

Concen: 53.120 ug/l

RT: 7.075 min Scan# 933

Delta R.T. -0.006 min

Lab File: VD078523.D




Acq: 22 Feb 2024 17:54


Tgt Ion: 77 Resp: 54050

Ion Ratio Lower Upper

77 100


97 23.7 13.2 39.6





#27
cis-1,2-Dichloroethene
Concen: 51.725 ug/l
RT: 7.081 min Scan# 91
Delta R.T. -0.000 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54




Instrument :
MSVOA_D
ClientSampleId :
20241101-COMPOSITE-35N-N-4-06MS



Tgt Ion: 96 Resp: 4201
Ion Ratio Lower Upper
96 100
61 130.3 0.0 239.2
98 67.3 0.0 128.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024



#28
Bromochloromethane
Concen: 62.848 ug/l
RT: 7.428 min Scan# 993
Delta R.T. -0.000 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54

Tgt Ion: 49 Resp: 30214
Ion Ratio Lower Upper
49 100
129 2.7 0.0 6.0
130 89.5 83.2 124.8





#29

Tetrahydrofuran

Concen: 387.150 ug/l

RT: 7.440 min Scan# 997

Delta R.T. -0.012 min

Lab File: VD078523.D


Acq: 22 Feb 2024 17:54

Instrument :

MSVOA_D

ClientSampleId :

20241101-COMPOSITE-35N-N-4-06MS





Tgt Ion: 42 Resp: 3593

Ion	Ratio	Lower	Upper
42	100		
72	51.4	46.5	69.7
71	47.8	44.5	66.7

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024



#30

Chloroform

Concen: 58.810 ug/l

RT: 7.593 min Scan# 1021




Delta R.T. -0.006 min


Lab File: VD078523.D

Acq: 22 Feb 2024 17:54

Tgt Ion: 83 Resp: 75525


Ion	Ratio	Lower	Upper
83	100		
85	64.8	53.4	80.2





#31
Cyclohexane
Concen: 29.020 ug/l
RT: 7.875 min Scan# 1069
Delta R.T. -0.006 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54




Instrument :
MSVOA_D
ClientSampleId :
20241101-COMPOSITE-35N-N-4-06MS



Tgt Ion: 56 Resp: 30639
Ion Ratio Lower Upper
56 100
69 33.3 28.6 43.0
84 97.7 84.8 127.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024



#32
1,1,1-Trichloroethane
Concen: 52.231 ug/l
RT: 7.793 min Scan# 1055
Delta R.T. -0.006 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54

Tgt Ion: 97 Resp: 57189
Ion Ratio Lower Upper
97 100
99 65.4 52.0 78.0
61 40.3 33.8 50.8





#33

1,2-Dichloroethane-d4

Concen: 69.816 ug/l

RT: 8.228 min Scan# 1129

Delta R.T. -0.006 min

Lab File: VD078523.D

Acq: 22 Feb 2024 17:54

Instrument :

MSVOA_D

ClientSampleId :

20241101-COMPOSITE-35N-N-4-06MS

Tgt Ion: 65 Resp: 4357

Ion Ratio Lower Upper

65 100





67 52.8 0.0 115.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 02/23/2024

Supervised By :Semsettin Yesilyurt 02/23/2024



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.775 min Scan# 1222

Delta R.T. -0.006 min

Lab File: VD078523.D

Acq: 22 Feb 2024 17:54




Tgt Ion:114 Resp: 103917


Ion Ratio Lower Upper

114 100

63 17.8 0.0 33.4

88 13.8 0.0 26.2





#35

Dibromofluoromethane

Concen: 65.667 ug/l

RT: 7.810 min Scan# 1058

Delta R.T. -0.000 min

Lab File: VD078523.D

Acq: 22 Feb 2024 17:54

Instrument :

MSVOA_D

ClientSampleId :





20241101-COMPOSITE-35N-N-4-06MS

Tgt Ion:	113	Resp:	4334
Ion Ratio	Lower	Upper	
113	100		
111	102.3	84.7	127.1
192	19.0	16.8	25.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024



#36

1,1-Dichloropropene

Concen: 42.012 ug/l




RT: 8.010 min Scan# 1092


Delta R.T. -0.000 min

Lab File: VD078523.D

Acq: 22 Feb 2024 17:54

Tgt Ion:	75	Resp:	43081
Ion Ratio	Lower	Upper	
75	100		
110	35.6	19.9	59.6
77	31.9	24.7	37.1





#37

Ethyl Acetate

Concen: 56.720 ug/l

RT: 7.169 min Scan# 949

Delta R.T. -0.000 min

Lab File: VD078523.D


Acq: 22 Feb 2024 17:54

Instrument :

MSVOA_D

ClientSampleId :

20241101-COMPOSITE-35N-N-4-06MS






Tgt Ion:	43	Resp:	20389
Ion Ratio	Lower	Upper	
43	100		
61	19.6	14.9	22.3
70	11.6	10.0	15.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024



#38

Carbon Tetrachloride

Concen: 25.309 ug/l




RT: 7.987 min Scan# 1088


Delta R.T. -0.012 min

Lab File: VD078523.D

Acq: 22 Feb 2024 17:54

Tgt Ion:	117	Resp:	25702
Ion Ratio	Lower	Upper	
117	100		
119	97.3	78.0	117.0
121	34.7	24.2	36.2





#39

Methylcyclohexane

Concen: 18.125 ug/l

RT: 9.269 min Scan# 11306

Delta R.T. -0.006 min

Lab File: VD078523.D

Acq: 22 Feb 2024 17:54

Instrument :

MSVOA_D

ClientSampleId :





20241101-COMPOSITE-35N-N-4-06MS

Tgt Ion: 83	Resp: 22065
Ion Ratio	Lower Upper
83	100
55	76.7 49.4 74.0
98	52.5 38.4 57.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024



#40

Benzene

Concen: 52.964 ug/l




RT: 8.251 min Scan# 1133


Delta R.T. -0.000 min

Lab File: VD078523.D

Acq: 22 Feb 2024 17:54


Tgt Ion: 78	Resp: 156985
Ion Ratio	Lower Upper
78	100
77	23.5 19.4 29.0





#41
Methacrylonitrile
Concen: 59.665 ug/l
RT: 7.399 min Scan# 911
Delta R.T. -0.006 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54




Instrument :
MSVOA_D
ClientSampleId :
20241101-COMPOSITE-35N-N-4-06MS



Tgt Ion: 41 Resp: 11162
Ion Ratio Lower Upper
41 100
39 63.1 34.0 51.0
67 112.1 77.0 115.4
52 43.8 27.9 41.9#




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024



#42
1,2-Dichloroethane
Concen: 61.073 ug/l
RT: 8.322 min Scan# 1145
Delta R.T. -0.006 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54

Tgt Ion: 62 Resp: 46839
Ion Ratio Lower Upper
62 100
98 11.2 0.0 25.4





#43

Isopropyl Acetate

Concen: 65.520 ug/l

RT: 8.363 min Scan# 1152

Delta R.T. -0.000 min

Lab File: VD078523.D

Acq: 22 Feb 2024 17:54

Instrument :

MSVOA_D

ClientSampleId :

20241101-COMPOSITE-35N-N-4-06MS





Tgt Ion: 43 Resp: 4361

Ion	Ratio	Lower	Upper
43	100		
61	20.9	31.4	47.0
87	17.4	14.9	22.3

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024



#44

Trichloroethene

Concen: 48.505 ug/l

RT: 9.028 min Scan# 1265




Delta R.T. -0.006 min


Lab File: VD078523.D

Acq: 22 Feb 2024 17:54

Tgt Ion:130 Resp: 36678

Ion	Ratio	Lower	Upper
130	100		
95	93.4	0.0	179.2





#45

1,2-Dichloropropane

Concen: 56.085 ug/l

RT: 9.304 min Scan# 11

Delta R.T. -0.000 min

Lab File: VD078523.D

Acq: 22 Feb 2024 17:54

Instrument :

MSVOA_D

ClientSampleId :

20241101-COMPOSITE-35N-N-4-06MS

Tgt Ion: 63 Resp: 4078

Ion Ratio Lower Upper

63 100





65 30.4 24.5 36.7

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 02/23/2024

Supervised By :Semsettin Yesilyurt 02/23/2024



#46

Dibromomethane

Concen: 62.245 ug/l

RT: 9.387 min Scan# 1326

Delta R.T. -0.012 min

Lab File: VD078523.D

Acq: 22 Feb 2024 17:54




Tgt Ion: 93 Resp: 25461


Ion Ratio Lower Upper

93 100

95 85.7 67.7 101.5


174 91.8 85.4 128.2





#47
Bromodichloromethane
Concen: 55.915 ug/l
RT: 9.587 min Scan# 1360
Delta R.T. -0.000 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54




Instrument : MSVOA_D
ClientSampleId : 20241101-COMPOSITE-35N-N-4-06MS



Tgt Ion: 83 Resp: 56650
Ion Ratio Lower Upper
83 100
85 60.9 50.2 75.4
127 8.0 6.3 9.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024



#48
Methyl methacrylate
Concen: 77.041 ug/l
RT: 9.381 min Scan# 1325
Delta R.T. -0.000 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54

Tgt Ion: 41 Resp: 27207
Ion Ratio Lower Upper
41 100
69 100.0 102.2 153.2#
39 60.0 41.4 62.2





#49

1,4-Dioxane

Concen: 1652.894 ug/l

RT: 9.387 min Scan# 1326

Delta R.T. -0.006 min

Lab File: VD078523.D


Acq: 22 Feb 2024 17:54

Instrument :

MSVOA_D

ClientSampleId :

20241101-COMPOSITE-35N-N-4-06MS



Tgt Ion: 88 Resp: 732




Ion	Ratio	Lower	Upper
88	100		
43	31.5	19.0	28.6
58	64.6	43.5	65.3

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 02/23/2024

Supervised By :Semsettin Yesilyurt 02/23/2024



#50

Toluene-d8

Concen: 53.660 ug/l

RT: 10.269 min Scan# 1476




Delta R.T. -0.000 min


Lab File: VD078523.D

Acq: 22 Feb 2024 17:54

Tgt Ion: 98 Resp: 141694


Ion	Ratio	Lower	Upper
98	100		
100	63.8	50.3	75.5





#51
4-Methyl-2-Pentanone
Concen: 370.925 ug/l
RT: 10.157 min Scan# 1457
Delta R.T. -0.006 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54




Instrument : MSVOA_D
ClientSampleId : 20241101-COMPOSITE-35N-N-4-06MS



Tgt Ion: 43 Resp: 131044
Ion Ratio Lower Upper
43 100
58 42.2 38.9 58.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024



#52
Toluene
Concen: 47.814 ug/l
RT: 10.334 min Scan# 1487
Delta R.T. -0.000 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54


Tgt Ion: 92 Resp: 89622
Ion Ratio Lower Upper
92 100
91 171.7 143.0 214.4





#53
t-1,3-Dichloropropene
Concen: 44.226 ug/l
RT: 10.551 min Scan# 11
Delta R.T. -0.000 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54




Instrument :
MSVOA_D
ClientSampleId :
20241101-COMPOSITE-35N-N-4-06MS



Tgt Ion: 75 Resp: 4088
Ion Ratio Lower Upper
75 100
77 31.6 25.3 37.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024



#54
cis-1,3-Dichloropropene
Concen: 39.153 ug/l
RT: 10.016 min Scan# 1433
Delta R.T. -0.000 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54

Tgt Ion: 75 Resp: 43890
Ion Ratio Lower Upper
75 100
77 31.7 25.5 38.3
39 36.7 22.1 33.1#





#55

1,1,2-Trichloroethane

Concen: 64.910 ug/l

RT: 10.734 min Scan# 1555

Delta R.T. -0.000 min

Lab File: VD078523.D

Acq: 22 Feb 2024 17:54

Instrument :

MSVOA_D

ClientSampleId :





20241101-COMPOSITE-35N-N-4-06MS

Tgt Ion: 97	Resp: 35750
Ion Ratio	Lower Upper
97 100	
83 93.1	65.8 98.6
85 55.4	45.7 68.5
99 66.4	46.2 69.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024



#56

Ethyl methacrylate

Concen: 54.225 ug/l




RT: 10.598 min Scan# 1532


Delta R.T. -0.000 min

Lab File: VD078523.D

Acq: 22 Feb 2024 17:54

Tgt Ion: 69	Resp: 27626		
Ion Ratio	Lower	Upper	
69 100			
41 55.1	39.1	58.7	
39 32.7	20.0	30.0	#





#57

1,3-Dichloropropane

Concen: 62.062 ug/l

RT: 10.881 min Scan# 1580

Delta R.T. -0.000 min

Lab File: VD078523.D

Acq: 22 Feb 2024 17:54

Instrument :

MSVOA_D

ClientSampleId :

20241101-COMPOSITE-35N-N-4-06MS

Tgt Ion: 76 Resp: 5815

Ion Ratio Lower Upper

76 100





78 34.4 26.4 39.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 02/23/2024

Supervised By :Semsettin Yesilyurt 02/23/2024



#58

2-Chloroethyl Vinyl ether

Concen: 374.990 ug/l

RT: 9.869 min Scan# 1408

Delta R.T. -0.000 min

Lab File: VD078523.D




Acq: 22 Feb 2024 17:54


Tgt Ion: 63 Resp: 112533

Ion Ratio Lower Upper

63 100


106 29.3 25.3 37.9





#59
2-Hexanone
Concen: 354.861 ug/l
RT: 10.922 min Scan# 1587
Delta R.T. -0.000 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54




Instrument : MSVOA_D
ClientSampleId : 20241101-COMPOSITE-35N-N-4-06MS



Tgt Ion: 43 Resp: 94210
Ion Ratio Lower Upper
43 100
58 60.4 34.1 102.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024



#60
Dibromochloromethane
Concen: 57.054 ug/l
RT: 11.069 min Scan# 1612
Delta R.T. -0.006 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54


Tgt Ion:129 Resp: 37814
Ion Ratio Lower Upper
129 100
127 79.4 37.1 111.5





#61
1,2-Dibromoethane
Concen: 61.216 ug/l
RT: 11.181 min Scan# 1107
Delta R.T. -0.000 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54




Instrument : MSVOA_D
ClientSampleId : 20241101-COMPOSITE-35N-N-4-06MS




Tgt Ion:107 Resp: 31658
Ion Ratio Lower Upper
107 100
109 96.7 75.6 113.4

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024




#62
4-Bromofluorobenzene
Concen: 42.098 ug/l
RT: 12.569 min Scan# 1867
Delta R.T. -0.006 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54



Tgt Ion: 95 Resp: 32349
Ion Ratio Lower Upper
95 100
174 92.6 0.0 201.8
176 88.1 0.0 191.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.581 min Scan# 1

Delta R.T. -0.006 min

Lab File: VD078523.D

Acq: 22 Feb 2024 17:54

Instrument :

MSVOA_D

ClientSampleId :

20241101-COMPOSITE-35N-N-4-06MS





Tgt Ion:117 Resp: 8225

Ion	Ratio	Lower	Upper
117	100		
82	48.4	38.3	57.5
119	28.8	25.0	37.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024



#64

Tetrachloroethene

Concen: 60.105 ug/l

RT: 10.810 min Scan# 1568




Delta R.T. -0.000 min


Lab File: VD078523.D

Acq: 22 Feb 2024 17:54

Tgt Ion:164 Resp: 31228

Ion	Ratio	Lower	Upper
164	100		
166	121.1	99.2	148.8
129	87.7	68.4	102.6
131	83.7	63.8	95.8









#65
Chlorobenzene
Concen: 51.528 ug/l
RT: 11.604 min Scan# 1703
Delta R.T. -0.006 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54

Instrument : MSVOA_D
ClientSampleId : 20241101-COMPOSITE-35N-N-4-06MS

Tgt Ion:112 Resp: 8521
Ion Ratio Lower Upper
112 100
114 28.0 25.1 37.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024



#66
1,1,1,2-Tetrachloroethane
Concen: 60.067 ug/l
RT: 11.681 min Scan# 1716
Delta R.T. -0.006 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54

Tgt Ion:131 Resp: 36504
Ion Ratio Lower Upper
131 100
133 91.4 48.5 145.5
119 65.6 35.6 106.9









#67
Ethyl Benzene
Concen: 43.511 ug/l
RT: 11.681 min Scan# 1716
Delta R.T. -0.006 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54

Instrument : MSVOA_D
ClientSampleId : 20241101-COMPOSITE-35N-N-4-06MS

Tgt Ion: 91 Resp: 12892
Ion Ratio Lower Upper
91 100
106 33.7 26.5 39.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024



#68
m/p-Xylenes
Concen: 86.036 ug/l
RT: 11.792 min Scan# 1735
Delta R.T. -0.000 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54


Tgt Ion:106 Resp: 99437
Ion Ratio Lower Upper
106 100
91 196.9 147.0 220.6





#69
o-Xylene
Concen: 45.774 ug/l
RT: 12.122 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54




Instrument : MSVOA_D
ClientSampleId : 20241101-COMPOSITE-35N-N-4-06MS



Tgt Ion:106 Resp: 48278
Ion Ratio Lower Upper
106 100
91 207.8 97.9 293.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024



#70
Styrene
Concen: 40.957 ug/l
RT: 12.134 min Scan# 1793
Delta R.T. -0.006 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54


Tgt Ion:104 Resp: 70501
Ion Ratio Lower Upper
104 100
78 45.2 33.2 49.8
103 58.0 43.6 65.4





#71
Bromoform
Concen: 64.248 ug/l
RT: 12.298 min Scan# 1821
Delta R.T. -0.000 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54




Instrument : MSVOA_D
ClientSampleId : 20241101-COMPOSITE-35N-N-4-06MS



Tgt Ion:173 Resp: 2055
Ion Ratio Lower Upper
173 100
175 49.6 24.6 73.8
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.516 min Scan# 2028
Delta R.T. -0.006 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54

Tgt Ion:152 Resp: 28038
Ion Ratio Lower Upper
152 100
115 56.3 27.1 81.3
150 169.0 0.0 353.6









#73
Isopropylbenzene
Concen: 50.909 ug/l
RT: 12.422 min Scan# 1842
Delta R.T. -0.000 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54

Instrument : MSVOA_D
ClientSampleId : 20241101-COMPOSITE-35N-N-4-06MS

Tgt Ion:105 Resp: 95848
Ion Ratio Lower Upper
105 100
120 26.6 14.3 42.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024



#74
N-ethyl acetate
Concen: 52.131 ug/l
RT: 12.234 min Scan# 1810
Delta R.T. -0.000 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54

Tgt Ion: 43 Resp: 18568
Ion Ratio Lower Upper
43 100
70 52.6 45.0 67.4
55 29.4 23.8 35.6
61 28.7 22.1 33.1





#75
1,1,2,2-Tetrachloroethane
Concen: 108.127 ug/l
RT: 12.675 min Scan# 1885
Delta R.T. -0.000 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54





Instrument : MSVOA_D
ClientSampleId : 20241101-COMPOSITE-35N-N-4-06MS

Tgt Ion: 83 Resp: 4038

Ion	Ratio	Lower	Upper
83	100		
131	8.4	5.1	15.3
85	63.6	32.6	97.8

Manual Integrations
APPROVED




Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024




#76
1,2,3-Trichloropropane
Concen: 107.405 ug/l m
RT: 12.722 min Scan# 1893
Delta R.T. -0.006 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54

Tgt Ion: 75 Resp: 25566


Ion	Ratio	Lower	Upper
75	100		
77	0.0	0.0	0.0





#77
Bromobenzene
Concen: 67.607 ug/l
RT: 12.698 min Scan# 1889
Delta R.T. -0.006 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54




Instrument : MSVOA_D
ClientSampleId : 20241101-COMPOSITE-35N-N-4-06MS



Tgt Ion:156 Resp: 29914
Ion Ratio Lower Upper
156 100
77 187.0 76.8 230.4
158 100.7 48.6 145.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024



#78
n-propylbenzene
Concen: 44.076 ug/l
RT: 12.763 min Scan# 1900
Delta R.T. -0.000 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54

Tgt Ion: 91 Resp: 103412
Ion Ratio Lower Upper
91 100
120 22.3 12.4 37.2









#79
2-Chlorotoluene
Concen: 52.731 ug/l
RT: 12.845 min Scan# 1915
Delta R.T. -0.006 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54

Instrument :
MSVOA_D
ClientSampleId :
20241101-COMPOSITE-35N-N-4-06MS

Tgt Ion: 91 Resp: 66264
Ion Ratio Lower Upper
91 100
126 35.7 19.4 58.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024



#80
1,3,5-Trimethylbenzene
Concen: 44.099 ug/l
RT: 12.904 min Scan# 1924
Delta R.T. -0.000 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54


Tgt Ion:105 Resp: 68942
Ion Ratio Lower Upper
105 100
120 51.8 26.4 79.2





#81
trans-1,4-Dichloro-2-butene
Concen: 41.218 ug/l
RT: 12.469 min Scan# 1850
Delta R.T. -0.000 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54




Instrument : MSVOA_D
ClientSampleId : 20241101-COMPOSITE-35N-N-4-06MS



Tgt Ion: 75 Resp: 4720
Ion Ratio Lower Upper
75 100
53 82.2 63.0 94.4
89 42.7 37.9 56.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024



#82
4-Chlorotoluene
Concen: 52.025 ug/l
RT: 12.945 min Scan# 1931
Delta R.T. -0.006 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54

Tgt Ion: 91 Resp: 70216
Ion Ratio Lower Upper
91 100
126 36.3 19.4 58.4









#83
tert-Butylbenzene
Concen: 39.859 ug/l
RT: 13.169 min Scan# 1969
Delta R.T. -0.000 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54

Instrument : MSVOA_D
ClientSampleId : 20241101-COMPOSITE-35N-N-4-06MS

Tgt Ion:119 Resp: 52552
Ion Ratio Lower Upper
119 100
91 59.9 29.3 87.8
134 24.7 14.1 42.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024



#84
1,2,4-Trimethylbenzene
Concen: 45.316 ug/l
RT: 13.210 min Scan# 1976
Delta R.T. -0.006 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54


Tgt Ion:105 Resp: 71900
Ion Ratio Lower Upper
105 100
120 47.8 24.3 72.9





#85
 sec-Butylbenzene
 Concen: 32.633 ug/l
 RT: 13.345 min Scan# 1999
 Delta R.T. -0.000 min
 Lab File: VD078523.D
 Acq: 22 Feb 2024 17:54




Instrument : MSVOA_D
 ClientSampleId : 20241101-COMPOSITE-35N-N-4-06MS




Tgt Ion:105 Resp: 66720
 Ion Ratio Lower Upper
 105 100
 134 18.6 10.9 32.9

Manual Integrations
 APPROVED



Reviewed By :Mahesh Dadoda 02/23/2024
 Supervised By :Semsettin Yesilyurt 02/23/2024




#86
 p-Isopropyltoluene
 Concen: 32.391 ug/l
 RT: 13.457 min Scan# 2018
 Delta R.T. -0.006 min
 Lab File: VD078523.D
 Acq: 22 Feb 2024 17:54



Tgt Ion:119 Resp: 56411
 Ion Ratio Lower Upper
 119 100
 134 26.6 14.2 42.6
 91 21.7 10.5 31.5









#87
1,3-Dichlorobenzene
Concen: 45.793 ug/l
RT: 13.457 min Scan# 2019
Delta R.T. -0.006 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54

Instrument : MSVOA_D
ClientSampleId : 20241101-COMPOSITE-35N-N-4-06MS

Tgt Ion:146 Resp: 43249
Ion Ratio Lower Upper
146 100
111 35.9 17.9 53.8
148 61.8 32.3 96.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024



#88
1,4-Dichlorobenzene
Concen: 45.860 ug/l
RT: 13.539 min Scan# 2032
Delta R.T. -0.000 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54

Tgt Ion:146 Resp: 42553
Ion Ratio Lower Upper
146 100
111 37.6 17.4 52.3
148 63.6 31.7 95.1









#89
n-Butylbenzene
Concen: 26.764 ug/l
RT: 13.786 min Scan# 2074
Delta R.T. -0.006 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54

Instrument : MSVOA_D
ClientSampleId : 20241101-COMPOSITE-35N-N-4-06MS

Tgt Ion: 91 Resp: 44928
Ion Ratio Lower Upper
91 100
92 46.9 26.5 79.5
134 24.6 13.8 41.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024



#90
Hexachloroethane
Concen: 12.365 ug/l
RT: 14.051 min Scan# 2119
Delta R.T. -0.006 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54

Tgt Ion:117 Resp: 4426
Ion Ratio Lower Upper
117 100
201 78.2 38.5 115.5





#91

1,2-Dichlorobenzene

Concen: 49.617 ug/l

RT: 13.834 min Scan# 2082

Delta R.T. -0.000 min

Lab File: VD078523.D

Acq: 22 Feb 2024 17:54

Instrument :

MSVOA_D

ClientSampleId :





20241101-COMPOSITE-35N-N-4-06MS

Tgt Ion:	146	Resp:	3934
Ion Ratio	Lower	Upper	
146	100		
111	39.6	18.3	54.8
148	64.2	31.6	94.7

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024



#92

1,2-Dibromo-3-Chloropropane

Concen: 105.027 ug/l




RT: 14.445 min Scan# 2186


Delta R.T. -0.006 min

Lab File: VD078523.D

Acq: 22 Feb 2024 17:54

Tgt Ion:	75	Resp:	5136
Ion Ratio	Lower	Upper	
75	100		
155	84.4	47.9	143.8
157	115.3	59.9	179.7





#93

1,2,4-Trichlorobenzene

Concen: 24.165 ug/l

RT: 15.098 min Scan# 21

Delta R.T. -0.006 min

Lab File: VD078523.D

Acq: 22 Feb 2024 17:54

Instrument :

MSVOA_D

ClientSampleId :





20241101-COMPOSITE-35N-N-4-06MS

Tgt Ion:180	Resp:	1126
Ion Ratio	Lower	Upper
180	100	
182	97.4	48.5 145.6
145	29.7	14.2 42.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024



#94

Hexachlorobutadiene

Concen: 16.562 ug/l




RT: 15.204 min Scan# 2315


Delta R.T. -0.000 min

Lab File: VD078523.D

Acq: 22 Feb 2024 17:54


Tgt Ion:225	Resp:	4029
Ion Ratio	Lower	Upper
225	100	
223	67.7	31.6 94.8
227	61.9	31.7 95.1





#95
Naphthalene
Concen: 43.113 ug/l
RT: 15.333 min Scan# 2337
Delta R.T. -0.000 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54




Instrument : MSVOA_D
ClientSampleId : 20241101-COMPOSITE-35N-N-4-06MS



Tgt Ion:128 Resp: 37268
Ion Ratio Lower Upper
128 100
127 12.5 10.5 15.7
129 10.0 8.7 13.1

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 02/23/2024
Supervised By :Semsettin Yesilyurt 02/23/2024



#96
1,2,3-Trichlorobenzene
Concen: 26.429 ug/l
RT: 15.522 min Scan# 2369
Delta R.T. -0.000 min
Lab File: VD078523.D
Acq: 22 Feb 2024 17:54

Tgt Ion:180 Resp: 10439
Ion Ratio Lower Upper
180 100
182 88.7 47.3 141.9
145 28.5 15.1 45.3

