

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022324\
 Data File : VD078534.D
 Acq On : 23 Feb 2024 14:22
 Operator : RP/MD
 Sample : P1538-09RE
 Misc : 5.00G/5ml/MSVOA_D/SOIL/B
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 20241102-AMENDED-COMPOSITE-35N-N-4-01RE

Quant Time: Feb 24 00:21:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021324S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 14 01:16:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	103527	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	202236	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	235829	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	106629	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	45472	40.622	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	81.240%
35) Dibromofluoromethane	7.804	113	28889	22.488	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	44.980%#
50) Toluene-d8	10.269	98	252083	49.054	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.100%
62) 4-Bromofluorobenzene	12.575	95	92929	62.141	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	124.280%
Target Compounds						
					Qvalue	
16) Acetone	4.028	43	23343	91.474	ug/l	90
17) Carbon Disulfide	4.269	76	136594	35.969	ug/l	98
25) 2-Butanone	7.099	43	5342	3.346	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022324\
Data File : VD078534.D
Acq On : 23 Feb 2024 14:22
Operator : RP/MD
Sample : P1538-09RE
Misc : 5.00G/5ml/MSVOA_D/SOIL/B
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
20241102-AMENDED-COMPOSITE-35N-N-4-01RE

Quant Time: Feb 24 00:21:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021324S.M
Quant Title : SW846 8260
QLast Update : Wed Feb 14 01:16:20 2024
Response via : Initial Calibration

