

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022324\
 Data File : VD078540.D
 Acq On : 23 Feb 2024 17:10
 Operator : RP/MD
 Sample : P1537-04
 Misc : 4.05G/5ml/MSVOA_D/SOIL/B
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 4

Quant Time: Feb 24 00:25:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021324S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 14 01:16:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	113049	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	212210	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.580	117	264784	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	135021	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	45760	37.436	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	74.880%
35) Dibromofluoromethane	7.804	113	65784	48.801	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.600%
50) Toluene-d8	10.269	98	278740	51.692	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	103.380%
62) 4-Bromofluorobenzene	12.575	95	108631	69.227	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	138.460%
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.798	84	16963	9.764	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022324\
Data File : VD078540.D
Acq On : 23 Feb 2024 17:10
Operator : RP/MD
Sample : P1537-04
Misc : 4.05G/5ml/MSVOA_D/SOIL/B
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
4

Quant Time: Feb 24 00:25:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021324S.M
Quant Title : SW846 8260
QLast Update : Wed Feb 14 01:16:20 2024
Response via : Initial Calibration

