



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022324\
Data File : VD078528.D
Acq On : 23 Feb 2024 08:40
Operator : RP/MD
Sample : BFB
Misc : 5.00G/5ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021324S.M
Title : SW846 8260
Last Update : Wed Feb 14 01:16:20 2024



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1859

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.3	15293	PASS
75	95	30	60	51.0	40387	PASS
95	95	100	100	100.0	79147	PASS
96	95	5	9	6.4	5083	PASS
173	174	0.00	2	1.4	945	PASS
174	95	50	100	86.0	68067	PASS
175	174	5	9	7.3	4994	PASS
176	174	95	101	98.2	66824	PASS
177	176	5	9	7.0	4700	PASS