

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD022420\
 Data File : VD065295.D
 Acq On : 24 Feb 2020 15:27
 Operator : VA/SY
 Sample : L1635-14
 Misc : 5.04G/5.00ml/MSVOA D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 LIED-TS001-02

Quant Time: Feb 25 02:47:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D021820S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:26:54 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.99	168	589702	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	976041	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	871518	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	378309	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.34	65	290579	54.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.54%	
35) Dibromofluoromethane	7.92	113	311843	52.40	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.80%	
50) Toluene-d8	10.34	98	1176784	51.72	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.44%	
62) 4-Bromofluorobenzene	12.64	95	343509	47.66	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.32%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD022420\
Data File : VD065295.D
Acq On : 24 Feb 2020 15:27
Operator : VA/SY
Sample : L1635-14
Misc : 5.04G/5.00ml/MSVOA D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
LIED-TS001-02

Quant Time: Feb 25 02:47:21 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D021820S.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:26:54 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.99 min Scan# 1031

Delta R.T. 0.00 min

Lab File: VD065295.D

Acq: 24 Feb 2020 15:27

Instrument :

MSVOA_D

ClientSampleId :


LIED-TS001-02

Tot Ion:168 Resp: 589702

Ion Ratio Lower Upper

168 100

99 51.6 41.4 62.2



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VDC

7.99

250000

200000



150000

100000

50000

0

Time-->



#33

1,2-Dichloroethane-d4

Concen: 54.774 ug/l

RT: 8.34 min Scan# 1091

Delta R.T. 0.00 min

Lab File: VD065295.D


Acq: 24 Feb 2020 15:27

Tgt Ion: 65 Resp: 290579

Ion Ratio Lower Upper

65 100

67 53.0 0.0 106.8



Abundance Ion 64.90 (64.60 to 65.60): VDC

Ion 66.90 (66.60 to 67.60): VDC

8.34


150000


100000

50000

0

Time-->





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.87 min Scan# 1181

Delta R.T. 0.00 min

Lab File: VD065295.D

Acq: 24 Feb 2020 15:27

Instrument :

MSVOA_D

ClientSampleId :

LIED-TS001-02


Tot Ion:114 Resp: 976041

Ion Ratio Lower Upper

114 100

63 20.1 0.0 39.4




88 15.3 0.0 30.4



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): V

Ion 87.90 (87.60 to 88.60): V



#35

Dibromofluoromethane

Concen: 52.399 ug/l

RT: 7.92 min Scan# 1020

Delta R.T. 0.00 min

Lab File: VD065295.D

Acq: 24 Feb 2020 15:27


Tgt Ion:113 Resp: 311843

Ion Ratio Lower Upper

113 100

111 101.1 82.9 124.3



192 22.0 17.3 25.9



Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 51.724 ug/l

RT: 10.34 min Scan# 1432

Delta R.T. 0.00 min

Lab File: VD065295.D

Acq: 24 Feb 2020 15:27

Instrument :

MSVOA_D

ClientSampleId :





LIED-TS001-02

Tot Ion: 98 Resp: 1176784

Ion Ratio Lower Upper

98 100

100 65.0 52.0 78.0



#62

4-Bromofluorobenzene

Concen: 47.660 ug/l

RT: 12.64 min Scan# 1822

Delta R.T. 0.00 min

Lab File: VD065295.D

Acq: 24 Feb 2020 15:27




Tgt Ion: 95 Resp: 343509


Ion Ratio Lower Upper

95 100

174 90.1 0.0 178.0

176 85.5 0.0 174.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.65 min Scan# 1654
Delta R.T. 0.00 min
Lab File: VD065295.D
Acq: 24 Feb 2020 15:27

Instrument :
MSVOA_D
ClientSampleId :
LIED-TS001-02

Tot Ion:117 Resp: 871518
Ion Ratio Lower Upper
117 100
82 53.4 42.7 64.1
119 32.6 25.5 38.3



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.58 min Scan# 1983
Delta R.T. 0.00 min
Lab File: VD065295.D
Acq: 24 Feb 2020 15:27

Tgt Ion:152 Resp: 378309
Ion Ratio Lower Upper
152 100
115 55.1 27.6 82.8
150 156.2 0.0 347.4

