


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD022420\
 Data File : VD065283.D
 Acq On : 24 Feb 2020 09:25
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D021820S.M
 Title : SW846 8260
 Last Update : Wed Feb 19 03:26:54 2020



AutoFind: Scans 1821, 1822, 1823; Background Corrected with Scan 1813

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.5	50002	PASS
75	95	30	60	48.4	130813	PASS
95	95	100	100	100.0	270357	PASS
96	95	5	9	7.1	19299	PASS
173	174	0.00	2	1.3	3012	PASS
174	95	50	100	86.6	234176	PASS
175	174	5	9	7.7	17991	PASS
176	174	95	101	96.3	225621	PASS
177	176	5	9	7.2	16193	PASS