

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022421\
 Data File : VD068434.D
 Acq On : 24 Feb 2021 13:46
 Operator : VA/SY
 Sample : M1460-01
 Misc : 3.20G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 HD-01-022221

Quant Time: Feb 25 00:10:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 12 13:30:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	376301	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	716356	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	658131	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	292936	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.338	65	206795	53.76	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 107.52%	
35) Dibromofluoromethane	7.914	113	214184	50.37	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 100.74%	
50) Toluene-d8	10.343	98	886953	50.33	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 100.66%	
62) 4-Bromofluorobenzene	12.632	95	284519	49.56	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 99.12%	
Target Compounds						Qvalue
95) Naphthalene	15.396	128	179558	18.39	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022421\
Data File : VD068434.D
Acq On : 24 Feb 2021 13:46
Operator : VA/SY
Sample : M1460-01
Misc : 3.20G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
HD-01-022221

Quant Time: Feb 25 00:10:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021221S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 12 13:30:51 2021
Response via : Initial Calibration

