

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022421\
 Data File : VD068436.D
 Acq On : 24 Feb 2021 15:10
 Operator : VA/SY
 Sample : M1469-01
 Misc : 7.33G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TR-08-022321

Quant Time: Feb 25 00:11:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 12 13:30:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	366157	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	694007	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	624214	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	237774	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.338	65	200141	53.47	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	106.94%
35) Dibromofluoromethane	7.920	113	208383	50.59	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.18%
50) Toluene-d8	10.344	98	863425	50.57	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.14%
62) 4-Bromofluorobenzene	12.632	95	254952	45.84	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	91.68%

Target Compounds	Qvalue
-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022421\
Data File : VD068436.D
Acq On : 24 Feb 2021 15:10
Operator : VA/SY
Sample : M1469-01
Misc : 7.33G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TR-08-022321

Quant Time: Feb 25 00:11:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021221S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 12 13:30:51 2021
Response via : Initial Calibration

