



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022421\
 Data File : VD068425.D
 Acq On : 24 Feb 2021 09:06
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D0212215.M
 Title : SW846 8260
 Last Update : Fri Feb 12 13:30:51 2021



AutoFind: Scans 1820, 1821, 1822; Background Corrected with Scan 1811

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.7	20693	PASS
75	95	30	60	49.3	57453	PASS
95	95	100	100	100.0	116613	PASS
96	95	5	9	6.7	7812	PASS
173	174	0.00	2	1.0	917	PASS
174	95	50	100	82.4	96112	PASS
175	174	5	9	7.4	7111	PASS
176	174	95	101	95.7	92027	PASS
177	176	5	9	7.2	6661	PASS