

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022522\
 Data File : VD072097.D
 Acq On : 25 Feb 2022 13:42
 Operator : VA/SY
 Sample : N1663-01
 Misc : 5.07G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CL-02-022422

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 02/28/2022
 Supervised By :Mahesh Dadoda 02/28/2022

Quant Time: Feb 26 05:41:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 19 04:05:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	42920	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	69094	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	74943	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	34875	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	36328	81.598	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 163.200%#	
35) Dibromofluoromethane	7.914	113	28497	68.074	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 136.140%	
50) Toluene-d8	10.338	98	86971	56.023	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 112.040%	
62) 4-Bromofluorobenzene	12.620	95	33150	58.133	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 116.260%	
Target Compounds					Qvalue	
16) Acetone	4.144	43	10852m	112.786	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022522\
Data File : VD072097.D
Acq On : 25 Feb 2022 13:42
Operator : VA/SY
Sample : N1663-01
Misc : 5.07G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CL-02-022422

Quant Time: Feb 26 05:41:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 19 04:05:38 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 02/28/2022
Supervised By : Mahesh Dadoda 02/28/2022

