

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022522\
 Data File : VD072099.D
 Acq On : 25 Feb 2022 14:39
 Operator : VA/SY
 Sample : N1655-02
 Misc : 5.18G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 REG-MID

Quant Time: Feb 26 05:42:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 19 04:05:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	103745	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	177188	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	192249	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	95884	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	86462	80.344	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 160.680%	
35) Dibromofluoromethane	7.909	113	68191	63.520	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 127.040%	
50) Toluene-d8	10.332	98	217476	54.627	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 109.260%	
62) 4-Bromofluorobenzene	12.620	95	86498	59.149	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 118.300%	
Target Compounds						
16) Acetone	4.156	43	6357	27.333	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022522\
Data File : VD072099.D
Acq On : 25 Feb 2022 14:39
Operator : VA/SY
Sample : N1655-02
Misc : 5.18G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
REG-MID

Quant Time: Feb 26 05:42:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 19 04:05:38 2022
Response via : Initial Calibration

