

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022522\
 Data File : VD072109.D
 Acq On : 25 Feb 2022 19:23
 Operator : VA/SY
 Sample : N1663-01RE
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CL-02-022422RE

Quant Time: Feb 26 05:44:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 19 04:05:38 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	139300	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	246145	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	268061	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	131298	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	107733	74.558	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	149.120%
35) Dibromofluoromethane	7.914	113	91265	61.198	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	122.400%
50) Toluene-d8	10.338	98	306593	55.437	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	110.880%
62) 4-Bromofluorobenzene	12.626	95	124211	61.143	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	122.280%
Target Compounds						Qvalue
16) Acetone	4.156	43	10239	32.788	ug/l	94
31) Cyclohexane	7.967	56	3723	1.459	ug/l #	64

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022522\
Data File : VD072109.D
Acq On : 25 Feb 2022 19:23
Operator : VA/SY
Sample : N1663-01RE
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CL-02-022422RE

Quant Time: Feb 26 05:44:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 19 04:05:38 2022
Response via : Initial Calibration

