

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022522\
 Data File : VD072110.D
 Acq On : 25 Feb 2022 19:51
 Operator : VA/SY
 Sample : N1670-01
 Misc : 5.04G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OR-03-022422

Quant Time: Feb 26 05:44:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 19 04:05:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	341938	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	600964	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	611943	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	295651	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	212427	59.891	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.780%
35) Dibromofluoromethane	7.914	113	200943	55.188	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.380%
50) Toluene-d8	10.338	98	746878	55.314	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	110.620%
62) 4-Bromofluorobenzene	12.626	95	293233	59.121	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	118.240%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022522\
Data File : VD072110.D
Acq On : 25 Feb 2022 19:51
Operator : VA/SY
Sample : N1670-01
Misc : 5.04G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OR-03-022422

Quant Time: Feb 26 05:44:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 19 04:05:38 2022
Response via : Initial Calibration





Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.979 min Scan# 1029

Delta R.T. 0.006 min

Lab File: VD072110.D

Acq: 25 Feb 2022 19:51

Instrument :

MSVOA_D

ClientSampleId :





OR-03-022422

Tgt Ion:168 Resp: 341938

Ion Ratio Lower Upper

168 100

99 58.7 47.1 70.7



1,2-Dichloroethane-d4

Concen: 59.891 ug/l

RT: 8.326 min Scan# 1089

Delta R.T. 0.000 min

Lab File: VD072110.D




Acq: 25 Feb 2022 19:51


Tgt Ion: 65 Resp: 212427

Ion Ratio Lower Upper

65 100

67 51.3 0.0 105.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.861 min Scan# 1180

Delta R.T. 0.000 min

Lab File: VD072110.D

Acq: 25 Feb 2022 19:51

Instrument :

MSVOA_D

ClientSampleId :

OR-03-022422





Tgt Ion:114 Resp: 600964

Ion Ratio Lower Upper

114 100

63 21.1 0.0 42.2

88 15.9 0.0 31.2



#35

Dibromofluoromethane

Concen: 55.188 ug/l

RT: 7.914 min Scan# 1019

Delta R.T. 0.006 min

Lab File: VD072110.D

Acq: 25 Feb 2022 19:51




Tgt Ion:113 Resp: 200943


Ion Ratio Lower Upper

113 100

111 101.1 83.0 124.4

192 19.1 15.3 22.9









#50
Toluene-d8
Concen: 55.314 ug/l
RT: 10.338 min Scan# 1430
Delta R.T. 0.006 min
Lab File: VD072110.D
Acq: 25 Feb 2022 19:51




Instrument :
MSVOA_D
ClientSampleId :
OR-03-022422


Tgt Ion: 98 Resp: 746878
Ion Ratio Lower Upper
98 100
100 65.4 53.8 80.6



#62
4-Bromofluorobenzene
Concen: 59.121 ug/l
RT: 12.626 min Scan# 1820
Delta R.T. 0.006 min
Lab File: VD072110.D
Acq: 25 Feb 2022 19:51

Tgt Ion: 95 Resp: 293233
Ion Ratio Lower Upper
95 100
174 78.7 0.0 155.0
176 75.3 0.0 148.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.638 min Scan# 117
Delta R.T. 0.000 min
Lab File: VD072110.D
Acq: 25 Feb 2022 19:51

Instrument :
MSVOA_D
ClientSampleId :
OR-03-022422

Tgt Ion:117 Resp: 611943
Ion Ratio Lower Upper
117 100
82 55.9 45.4 68.0
119 32.7 25.4 38.2



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.561 min Scan# 1979
Delta R.T. 0.000 min
Lab File: VD072110.D
Acq: 25 Feb 2022 19:51

Tgt Ion:152 Resp: 295651
Ion Ratio Lower Upper
152 100
115 60.3 29.9 89.7
150 158.4 0.0 347.6

