


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD022620\
 Data File : VD065317.D
 Acq On : 26 Feb 2020 09:42
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D021820S.M
 Title : SW846 8260
 Last Update : Wed Feb 19 03:26:54 2020



AutoFind: Scans 1822, 1823, 1824; Background Corrected with Scan 1812

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.1	50813	PASS
75	95	30	60	48.7	136938	PASS
95	95	100	100	100.0	281045	PASS
96	95	5	9	6.4	17901	PASS
173	174	0.00	2	1.7	4343	PASS
174	95	50	100	90.8	255317	PASS
175	174	5	9	7.5	19098	PASS
176	174	95	101	95.6	244117	PASS
177	176	5	9	7.0	17042	PASS