

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022621\
 Data File : VD068460.D
 Acq On : 26 Feb 2021 16:57
 Operator : VA/SY
 Sample : M1477-01
 Misc : 5.07G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 72-11995

Quant Time: Feb 26 21:48:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 26 14:04:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	364480	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	676286	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	629037	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	286289	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	188078	51.65	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.30%
35) Dibromofluoromethane	7.920	113	111102	27.27	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	54.54%
50) Toluene-d8	10.338	98	851808	52.49	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.98%
62) 4-Bromofluorobenzene	12.632	95	269886	51.61	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	103.22%
Target Compounds						
16) Acetone	4.162	43	17116	29.59	ug/l	96
20) Methylene Chloride	4.967	84	55242	9.39	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022621\
Data File : VD068460.D
Acq On : 26 Feb 2021 16:57
Operator : VA/SY
Sample : M1477-01
Misc : 5.07G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
72-11995

Quant Time: Feb 26 21:48:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 26 14:04:14 2021
Response via : Initial Calibration

