

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022621\
 Data File : VD068462.D
 Acq On : 26 Feb 2021 17:53
 Operator : VA/SY
 Sample : M1491-01RE
 Misc : 6.30G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SA-01-022521RE

Quant Time: Feb 26 21:49:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 26 14:04:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	100461	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	190873	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	192977	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	85913	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.338	65	72889	72.62	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 145.24%	
35) Dibromofluoromethane	7.920	113	64215	55.85	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 111.70%	
50) Toluene-d8	10.338	98	237520	51.86	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 103.72%	
62) 4-Bromofluorobenzene	12.632	95	83027	56.25	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 112.50%	
Target Compounds						
16) Acetone	4.162	43	12198	76.51	ug/l	99
52) Toluene	10.414	92	5045	1.52	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022621\
Data File : VD068462.D
Acq On : 26 Feb 2021 17:53
Operator : VA/SY
Sample : M1491-01RE
Misc : 6.30G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SA-01-022521RE

Quant Time: Feb 26 21:49:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 26 14:04:14 2021
Response via : Initial Calibration

