

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022621\
 Data File : VD068465.D
 Acq On : 26 Feb 2021 19:17
 Operator : VA/SY
 Sample : M1518-02
 Misc : 7.68G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-1D

Quant Time: Feb 26 21:50:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 26 14:04:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	418030	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	771425	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	750167	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	347491	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	232483	55.66	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.32%
35) Dibromofluoromethane	7.920	113	243563	52.41	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.82%
50) Toluene-d8	10.344	98	965244	52.15	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.30%
62) 4-Bromofluorobenzene	12.632	95	324884	54.46	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	108.92%
Target Compounds						
16) Acetone	4.156	43	10386	15.65	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022621\
Data File : VD068465.D
Acq On : 26 Feb 2021 19:17
Operator : VA/SY
Sample : M1518-02
Misc : 7.68G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-1D

Quant Time: Feb 26 21:50:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 26 14:04:14 2021
Response via : Initial Calibration

