

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022621\
 Data File : VD068468.D
 Acq On : 26 Feb 2021 20:41
 Operator : VA/SY
 Sample : M1518-05
 Misc : 14.60G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-6

Quant Time: Feb 26 21:51:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 26 14:04:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	383187	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	682740	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	642717	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	280133	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	206738	54.00	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 108.00%		
35) Dibromofluoromethane	7.920	113	215965	52.51	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 105.02%		
50) Toluene-d8	10.344	98	864277	52.76	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	= 105.52%		
62) 4-Bromofluorobenzene	12.632	95	280057	53.05	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	= 106.10%		
Target Compounds						
64) Tetrachloroethene	10.879	164	8489	2.02	ug/l	Qvalue # 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022621\
Data File : VD068468.D
Acq On : 26 Feb 2021 20:41
Operator : VA/SY
Sample : M1518-05
Misc : 14.60G/5.00ml/MSVOA_D/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-6

Quant Time: Feb 26 21:51:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 26 14:04:14 2021
Response via : Initial Calibration

