



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_D\Data\VD022624\  
Data File : VD078544.D  
Acq On : 26 Feb 2024 08:46  
Operator : RP/MD  
Sample : BFB  
Misc : 5.00G/5ml/MSVOA\_D/SOIL  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_D  
ClientSampleId :  
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA\_D\Method\82D021324S.M  
Title : SW846 8260  
Last Update : Wed Feb 14 01:16:20 2024



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1859

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 50             | 95              | 15              | 40              | 18.1         | 15724      | PASS                |
| 75             | 95              | 30              | 60              | 50.4         | 43803      | PASS                |
| 95             | 95              | 100             | 100             | 100.0        | 86963      | PASS                |
| 96             | 95              | 5               | 9               | 6.9          | 5959       | PASS                |
| 173            | 174             | 0.00            | 2               | 1.3          | 1087       | PASS                |
| 174            | 95              | 50              | 100             | 93.7         | 81485      | PASS                |
| 175            | 174             | 5               | 9               | 8.6          | 6986       | PASS                |
| 176            | 174             | 95              | 101             | 96.7         | 78779      | PASS                |
| 177            | 176             | 5               | 9               | 6.5          | 5129       | PASS                |