

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022822\
 Data File : VD072120.D
 Acq On : 28 Feb 2022 14:03
 Operator : VA/SY
 Sample : N1655-07RE
 Misc : 6.89G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 REMOTE-1RE

Quant Time: Mar 01 08:12:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 19 04:05:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	45651	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	74463	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	78875	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	38059	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	34414	72.675	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	145.340%
35) Dibromofluoromethane	7.914	113	28695	63.604	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	127.200%
50) Toluene-d8	10.332	98	88821	53.089	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.180%
62) 4-Bromofluorobenzene	12.626	95	34851	56.709	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	113.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022822\
Data File : VD072120.D
Acq On : 28 Feb 2022 14:03
Operator : VA/SY
Sample : N1655-07RE
Misc : 6.89G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
REMOTE-1RE

Quant Time: Mar 01 08:12:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 19 04:05:38 2022
Response via : Initial Calibration

