

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022822\
 Data File : VD072121.D
 Acq On : 28 Feb 2022 14:31
 Operator : VA/SY
 Sample : N1678-01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OK-02-022522

Quant Time: Mar 01 08:12:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 19 04:05:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	537265	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	936485	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	900493	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	416241	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	262939	47.181	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	94.360%
35) Dibromofluoromethane	7.914	113	286159	50.434	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.860%
50) Toluene-d8	10.332	98	1145968	54.463	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.920%
62) 4-Bromofluorobenzene	12.626	95	407895	52.775	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	105.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022822\
Data File : VD072121.D
Acq On : 28 Feb 2022 14:31
Operator : VA/SY
Sample : N1678-01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OK-02-022522

Quant Time: Mar 01 08:12:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 19 04:05:38 2022
Response via : Initial Calibration

