

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022822\
 Data File : VD072124.D
 Acq On : 28 Feb 2022 15:57
 Operator : VA/SY
 Sample : VIBLK
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: Mar 01 08:13:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 19 04:05:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	532689	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	927444	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	907442	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	432271	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	272887	49.386	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.780%
35) Dibromofluoromethane	7.908	113	283101	50.382	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.760%
50) Toluene-d8	10.338	98	1128121	54.138	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.280%
62) 4-Bromofluorobenzene	12.626	95	421622	55.082	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	110.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022822\
Data File : VD072124.D
Acq On : 28 Feb 2022 15:57
Operator : VA/SY
Sample : VIBLK
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Quant Time: Mar 01 08:13:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 19 04:05:38 2022
Response via : Initial Calibration

