

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022822\
 Data File : VD072127.D
 Acq On : 28 Feb 2022 17:22
 Operator : VA/SY
 Sample : N1711-01
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 NB-07-022822

Quant Time: Mar 01 08:13:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 19 04:05:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	280174	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	495798	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	516277	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	256450	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	185644	63.878	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 127.760%	
35) Dibromofluoromethane	7.914	113	170535	56.771	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 113.540%	
50) Toluene-d8	10.337	98	618616	55.533	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 111.060%	
62) 4-Bromofluorobenzene	12.626	95	243308	59.461	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 118.920%	
Target Compounds						
16) Acetone	4.161	43	8734	13.906	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD022822\
Data File : VD072127.D
Acq On : 28 Feb 2022 17:22
Operator : VA/SY
Sample : N1711-01
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
NB-07-022822

Quant Time: Mar 01 08:13:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 19 04:05:38 2022
Response via : Initial Calibration

