

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030121\
 Data File : VD068479.D
 Acq On : 01 Mar 2021 14:50
 Operator : VA/SY
 Sample : M1526-01
 Misc : 7.20G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 STOCK-PILE

Quant Time: Mar 01 20:47:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 26 14:04:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	28439	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	52867	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	56603	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	25854	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.338	65	23681	83.34	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 166.68%#	
35) Dibromofluoromethane	7.920	113	17533	55.05	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 110.10%	
50) Toluene-d8	10.343	98	65552	51.68	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 103.36%	
62) 4-Bromofluorobenzene	12.632	95	23310	57.02	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 114.04%	
Target Compounds						
16) Acetone	4.161	43	20850	461.96	ug/l	93
52) Toluene	10.408	92	20024	21.75	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030121\
Data File : VD068479.D
Acq On : 01 Mar 2021 14:50
Operator : VA/SY
Sample : M1526-01
Misc : 7.20G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
STOCK-PILE

Quant Time: Mar 01 20:47:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 26 14:04:14 2021
Response via : Initial Calibration

