

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030121\
 Data File : VD068482.D
 Acq On : 01 Mar 2021 16:42
 Operator : VA/SY
 Sample : M1526-01
 Misc : 7.67G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 STOCK-PILE

Quant Time: Mar 01 20:48:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 26 14:04:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	336569	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	625211	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	607307	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	274318	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	194863	57.95	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 115.90%			
35) Dibromofluoromethane	7.914	113	195679	51.96	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 103.92%			
50) Toluene-d8	10.344	98	775523	51.70	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 103.40%			
62) 4-Bromofluorobenzene	12.632	95	260253	53.83	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 107.66%			
Target Compounds						
16) Acetone	4.162	43	16824	31.50	ug/l	97
52) Toluene	10.408	92	17307	1.59	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030121\
Data File : VD068482.D
Acq On : 01 Mar 2021 16:42
Operator : VA/SY
Sample : M1526-01
Misc : 7.67G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
STOCK-PILE

Quant Time: Mar 01 20:48:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 26 14:04:14 2021
Response via : Initial Calibration

