

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030122\
 Data File : VD072138.D
 Acq On : 01 Mar 2022 14:33
 Operator : VA/SY
 Sample : VIBLK
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: Mar 02 06:09:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 19 04:05:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	520178	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	898049	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	866141	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	413298	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	245737	45.542	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	91.080%
35) Dibromofluoromethane	7.914	113	268786	49.400	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.800%
50) Toluene-d8	10.338	98	1110380	55.031	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	110.060%
62) 4-Bromofluorobenzene	12.626	95	398355	53.746	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	107.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030122\
Data File : VD072138.D
Acq On : 01 Mar 2022 14:33
Operator : VA/SY
Sample : VIBLK
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Quant Time: Mar 02 06:09:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D021822S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 19 04:05:38 2022
Response via : Initial Calibration

