

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030124\
 Data File : VD078579.D
 Acq On : 01 Mar 2024 15:00
 Operator : RP/MD
 Sample : VIBLK
 Misc : 5.00G/5ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: Mar 02 03:58:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D030124S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 02 03:50:17 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	146435	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	270956	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	323537	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	170495	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	52762	36.348	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	72.700%
35) Dibromofluoromethane	7.804	113	78975	44.807	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	89.620%
50) Toluene-d8	10.269	98	339134	49.516	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.040%
62) 4-Bromofluorobenzene	12.575	95	132047	67.503	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	135.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030124\
Data File : VD078579.D
Acq On : 01 Mar 2024 15:00
Operator : RP/MD
Sample : VIBLK
Misc : 5.00G/5ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Quant Time: Mar 02 03:58:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D030124S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 02 03:50:17 2024
Response via : Initial Calibration

