

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030221\
 Data File : VD068501.D
 Acq On : 02 Mar 2021 17:34
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 02 23:59:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 26 14:04:14 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	111	0.00
2 T	Dichlorodifluoromethane	0.485	0.418	13.8	95	0.00
3 P	Chloromethane	0.613	0.516	15.8	96	0.00
4 C	Vinyl Chloride	0.728	0.625	14.1#	95	0.00
5 T	Bromomethane	0.518	0.444	14.3	101	0.00
6 T	Chloroethane	0.471	0.417	11.5	101	0.00
7 T	Trichlorofluoromethane	0.831	0.772	7.1	104	0.00
8 T	Diethyl Ether	0.247	0.235	4.9	106	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.506	0.471	6.9	105	0.00
10 T	Methyl Iodide	0.537	0.522	2.8	98	0.00
11 T	Tert butyl alcohol	0.027	0.024	11.1	111	0.00
12 CM	1,1-Dichloroethene	0.502	0.456	9.2#	101	0.00
13 T	Acrolein	0.038	0.038	0.0	107	0.00
14 T	Allyl chloride	0.648	0.532	17.9	91	0.00
15 T	Acrylonitrile	0.104	0.096	7.7	103	0.00
16 T	Acetone	0.079	0.065	17.7	88	0.00
17 T	Carbon Disulfide	1.645	1.409	14.3	95	0.00
18 T	Methyl Acetate	0.205	0.194	5.4	110	0.00
19 T	Methyl tert-butyl Ether	1.089	1.052	3.4	110	0.00
20 T	Methylene Chloride	0.673	0.548	18.6	101	0.00
21 T	trans-1,2-Dichloroethene	0.600	0.554	7.7	104	0.00
22 T	Diisopropyl ether	1.428	1.252	12.3	97	0.00
23 T	Vinyl Acetate	0.741	0.671	9.4	99	0.00
24 P	1,1-Dichloroethane	0.979	0.870	11.1	99	0.00
25 T	2-Butanone	0.114	0.102	10.5	101	0.00
26 T	2,2-Dichloropropane	0.815	0.724	11.2	99	0.00
27 T	cis-1,2-Dichloroethene	0.658	0.618	6.1	105	0.00
28 T	Bromochloromethane	0.382	0.344	9.9	100	0.00
29 T	Tetrahydrofuran	0.074	0.067	9.5	102	0.00
30 C	Chloroform	1.033	0.963	6.8#	105	0.00
31 T	Cyclohexane	0.909	0.735	19.1	94	0.00
32 T	1,1,1-Trichloroethane	0.882	0.825	6.5	105	0.00
33 S	1,2-Dichloroethane-d4	0.500	0.464	7.2	104	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	108	0.00
35 S	Dibromofluoromethane	0.301	0.299	0.7	108	0.00
36 T	1,1-Dichloropropene	0.475	0.426	10.3	97	0.00
37 T	Ethyl Acetate	0.158	0.147	7.0	106	0.00
38 T	Carbon Tetrachloride	0.447	0.427	4.5	105	0.00
39 T	Methylcyclohexane	0.586	0.537	8.4	97	0.00
40 TM	Benzene	1.363	1.249	8.4	100	0.00
41 T	Methacrylonitrile	0.092	0.093	-1.1	119	0.01
42 TM	1,2-Dichloroethane	0.358	0.346	3.4	104	0.00
43 T	Isopropyl Acetate	0.290	0.279	3.8	103	0.00
44 TM	Trichloroethene	0.374	0.361	3.5	106	0.00
45 C	1,2-Dichloropropane	0.326	0.301	7.7#	99	0.00
46 T	Dibromomethane	0.177	0.178	-0.6	113	0.00
47 T	Bromodichloromethane	0.459	0.450	2.0	107	0.00
48 T	Methyl methacrylate	0.154	0.130	15.6	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030221\
 Data File : VD068501.D
 Acq On : 02 Mar 2021 17:34
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 02 23:59:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 26 14:04:14 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	110	0.00
50 S	Toluene-d8	1.200	1.126	6.2	103	0.00
51 T	4-Methyl-2-Pentanone	0.152	0.146	3.9	103	0.00
52 CM	Toluene	0.871	0.837	3.9#	104	0.00
53 T	t-1,3-Dichloropropene	0.424	0.414	2.4	108	0.00
54 T	cis-1,3-Dichloropropene	0.517	0.487	5.8	102	0.00
55 T	1,1,2-Trichloroethane	0.245	0.245	0.0	109	0.00
56 T	Ethyl methacrylate	0.282	0.294	-4.3	110	0.00
57 T	1,3-Dichloropropane	0.428	0.415	3.0	106	0.00
58 T	2-Chloroethyl Vinyl ether	0.141	0.151	-7.1	116	0.00
59 T	2-Hexanone	0.101	0.099	2.0	103	0.00
60 T	Dibromochloromethane	0.304	0.324	-6.6	115	0.00
61 T	1,2-Dibromoethane	0.240	0.245	-2.1	111	0.00
62 S	4-Bromofluorobenzene	0.387	0.381	1.6	107	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	109	0.00
64 T	Tetrachloroethene	0.327	0.316	3.4	109	0.00
65 PM	Chlorobenzene	1.018	0.976	4.1	107	0.00
66 T	1,1,1,2-Tetrachloroethane	0.354	0.353	0.3	111	0.00
67 C	Ethyl Benzene	1.818	1.713	5.8#	103	0.00
68 T	m/p-Xylenes	0.690	0.655	5.1	103	0.00
69 T	o-Xylene	0.636	0.617	3.0	104	0.00
70 T	Styrene	1.087	1.063	2.2	105	0.00
71 P	Bromoform	0.180	0.190	-5.6	115	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	113	0.00
73 T	Isopropylbenzene	3.708	3.381	8.8	103	0.00
74 T	N-amyl acetate	0.632	0.566	10.4	100	0.00
75 P	1,1,2,2-Tetrachloroethane	0.638	0.612	4.1	113	0.00
76 T	1,2,3-Trichloropropane	0.444	0.423	4.7	127	0.00
77 T	Bromobenzene	0.846	0.816	3.5	112	0.00
78 T	n-propylbenzene	4.431	3.964	10.5	100	0.00
79 T	2-Chlorotoluene	2.459	2.182	11.3	100	0.00
80 T	1,3,5-Trimethylbenzene	3.039	2.808	7.6	103	0.00
81 T	trans-1,4-Dichloro-2-butene	0.190	0.174	8.4	102	0.00
82 T	4-Chlorotoluene	2.588	2.300	11.1	100	0.00
83 T	tert-Butylbenzene	2.588	2.423	6.4	104	0.00
84 T	1,2,4-Trimethylbenzene	3.044	2.818	7.4	103	0.00
85 T	sec-Butylbenzene	3.656	3.301	9.7	101	0.00
86 T	p-Isopropyltoluene	3.318	3.068	7.5	102	0.00
87 T	1,3-Dichlorobenzene	1.635	1.535	6.1	105	0.00
88 T	1,4-Dichlorobenzene	1.619	1.514	6.5	107	0.00
89 T	n-Butylbenzene	3.201	2.899	9.4	100	0.00
90 T	Hexachloroethane	0.660	0.596	9.7	98	0.00
91 T	1,2-Dichlorobenzene	1.402	1.419	-1.2	115	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.112	0.102	8.9	107	0.00
93 T	1,2,4-Trichlorobenzene	1.083	1.048	3.2	108	0.00
94 T	Hexachlorobutadiene	0.601	0.568	5.5	105	0.00
95 T	Naphthalene	1.961	1.981	-1.0	113	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030221\
Data File : VD068501.D
Acq On : 02 Mar 2021 17:34
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Mar 02 23:59:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 26 14:04:14 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.937	0.913	2.6	111	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6