


Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030221\
 Data File : VD068484.D
 Acq On : 02 Mar 2021 08:34
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
 Title : SW846 8260
 Last Update : Fri Feb 26 14:04:14 2021



AutoFind: Scans 1820, 1821, 1822; Background Corrected with Scan 1812

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.5	13240	PASS
75	95	30	60	48.3	36419	PASS
95	95	100	100	100.0	75464	PASS
96	95	5	9	6.4	4845	PASS
173	174	0.00	2	1.0	681	PASS
174	95	50	100	87.8	66264	PASS
175	174	5	9	7.4	4902	PASS
176	174	95	101	95.4	63197	PASS
177	176	5	9	6.6	4173	PASS